

CORRECTED VERSION

(19) World Intellectual Property Organization
International Bureau



(43) International Publication Date
27 July 2000 (27.07.2000)


PCT

(10) International Publication Number
WO 00/42910 A1

- (51) International Patent Classification⁷: A61B 5/00 (72) Inventor; and
(21) International Application Number: PCT/US00/01673 (75) Inventor/Applicant (for US only): FULGHUM, Stephen, F., Jr. [US/US]; 9 Driftwood Road, Marblehead, MA 01945-1250 (US).
(22) International Filing Date: 26 January 2000 (26.01.2000) (74) Agents: HOOVER, Thomas, O. et al.; Hamilton, Brook, Smith & Reynolds, P.C., Two Militia Drive, Lexington, MA 02421 (US).
(25) Filing Language: English
(26) Publication Language: English
(30) Priority Data:
09/238,664 26 January 1999 (26.01.1999) US
09/362,806 28 July 1999 (28.07.1999) US
(63) Related by continuation (CON) or continuation-in-part (CIP) to earlier application:
US 09/362,806 (CIP)
Filed on 28 July 1999 (28.07.1999)
(71) Applicant (for all designated States except US): NEW-TON LABORATORIES, INC. [US/US]; Suite 4850, 800 West Cummings Park, Woburn, MA 01801 (US).
(81) Designated States (national): AE, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CR, CU, CZ, DE, DK, DM, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW.
(84) Designated States (regional): ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU,

[Continued on next page]

(54) Title: AUTOFLUORESCENCE IMAGING SYSTEM FOR ENDOSCOPY



(57) Abstract: A system and method for imaging tissue autofluorescence through a video endoscope is described, comprising a light source capable of providing both ultraviolet light capable of inducing tissue autofluorescence and visible light which induces little or no autofluorescence, an optical system to deliver both wavelength bands to the tissue with the same apparent spatial and angular intensity distribution, a means for digitally acquiring the resulting, visible fluorescence and visible reflectance images using a single imaging detector at the distal tip of the endoscope and a means for digitally processing said images to generate a final, false-color image for display which indicates regions of tissue dysplasia. This system can either be added on to an existing video endoscope or integrated into its structure. The combined system can be electronically switched between normal white light imaging and fluorescence imaging.

WO 00/42910 A1



MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).

(15) Information about Correction:

see PCT Gazette No. 43/2001 of 25 October 2001, Section II

Published:

— *with international search report*

(48) Date of publication of this corrected version:

25 October 2001

For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.

AUTOFLUORESCENCE IMAGING SYSTEM FOR ENDOSCOPY

GOVERNMENT SUPPORT

This invention was supported in whole or in part by a grant from the National Institute of Health, Grant No. R44CA72626, "Imaging Spectrofluorimeter
5 for Colonic Dysplasia". The Government has certain rights in the invention.

RELATED APPLICATION

This is a continuation-in-part application of U.S. Application No. 09/362,806, filed July 28, 1999 which is a continuation-in-part application of U.S. Application No. 09/238,664 filed on January 26, 1999, the entire contents of the
10 above applications being is incorporated herein by reference.

BACKGROUND OF THE INVENTION

Advanced stage cancers are difficult to treat effectively, so it is important to detect them in their premalignant stages, such as dysplasia or carcinoma in situ. Currently, the most widely used method for early detection uses visual inspection
15 through endoscopes, which relies upon the recognition of the gross architectural changes associated with dysplasia. Visual inspection is less effective in detecting the superficial lesions of flat dysplasia, such as ulcerative colitis and Barrett's esophagus. In these cases, surveillance requires the selection of representative sites for biopsy and subsequent histological analysis. Only a very small fraction of a large
20 surface, such as the colon, can be surveyed in this manner and small areas of dysplasia may go undetected. More efficient methods for detecting flat dysplasia would provide a significant means of reducing cancer morbidity, mortality, and costs.

A promising technique for detecting dysplasia during endoscopy involves illuminating the tissue with light of an appropriate wavelength and observing the resulting fluorescence. Tissue fluorescence occurs at longer wavelengths than the excitation illumination and is typically much weaker, so that spectroscopic techniques are generally required for its detection. Diagnostic methods which use fluorescence information can generally be divided into two groups. The first group of methods observes fluorescence from medications administered to the patient which have accumulated in tumor tissues. The second group of methods observes endogenous fluorescence or autofluorescence arising from substances natural to the tissue itself which change their relative concentrations when the tissue becomes dysplastic. Of the two general fluorescence methods, the first method, requiring a prior application of a drug, is the more invasive. The application of these drugs takes additional time and has the potential of causing adverse side effects. Methods based on autofluorescence detection are less invasive and more suited to endoscopy for screening purposes.

Normal colon tissue, for example, when illuminated with ultraviolet light at 370 nm, exhibits a broad, blue fluorescence with a peak at 450 nm, as shown in Figure 1. This fluorescence is due to collagen, the primary protein of connective tissue, which is found within the thin mucosal layer and which is the dominant component of the submucosal layer. The fluorescence of dysplastic colon tissue, due to changes in its structure and chemistry, is typically $\frac{1}{2}$ to $\frac{1}{3}$ as intense given the same illumination. This reduction in visible blue-green autofluorescence, produced by ultraviolet to violet excitation light, has been identified as a primary indicator of dysplastic tissue. An increase in the relative fluorescence at 680 nm compared to 600 nm is a secondary indicator of dysplasia.

Diagnostic instruments which detect autofluorescence can be divided into two general groups. The first subdivision includes instruments which utilize fiberoptic probes to perform essentially point measurements on the tissue. The second subdivision includes instruments which produce detailed, two dimensional images. Point detection instruments have the advantage of providing more complete

spectral information on the tissue, but are too slow for routine screening of large tissue areas and may miss small regions of dysplasia. Fluorescence imaging endoscopes are more appropriate for screening large tissue areas such as the colon.

Fluorescence imaging systems designed to sense concentrations of
5 fluorescent markers that have been applied to the region of interest are optimized for the measurement of relatively high fluorescence levels, do not describe the additional instrumental features required to measure inherently weak autofluorescence. In particular, they do not describe a method for providing sufficient out-of-band filtering for the excitation illumination that would allow the
10 effective measurement of autofluorescence, which typically is reduced in intensity, compared to the excitation intensity, by a factor of 1000 or more.

Fluorescence imaging endoscopes which are specifically designed to measure autofluorescence can be further divided into groups based on the excitation wavelength chosen and the method for quantifying the degree to which the
15 autofluorescence is reduced. These design choices have a direct bearing on commercial considerations for the instrument since they affect, for example, the number of imaging devices required, the opto-mechanical complexity of the instrument, and the handling characteristics of the instrument in actual use.

Existing fluorescence imaging endoscope systems use visible blue light near
20 440 nm for the excitation wavelength, resulting in fluorescence peaks around 500 nm. These instruments use Helium-Cadmium lasers at 442 nm as a bright, easily controlled source for the excitation light. The high cost of Helium-Cadmium lasers makes them impractical as sources for commercial instruments.

Multiple cameras and mechanically switched optical components and/or
25 filters require the use of endoscopes based on coherent imaging fiber bundles, so that the cameras and filters can be located at the proximal end of the endoscope where there is room for them. The coherent fiber imaging bundles introduce significant light losses and do not provide as sharp an image as those now available with video endoscopes.

SUMMARY OF THE INVENTION

The present invention relates to imaging endoscopes and in particular to an endoscope system and method for imaging autofluorescence from epithelial tissue to highlight regions of dysplasia. The system for detecting dysplastic tissue uses the autofluorescence of mucosal tissue such as that found in the colon, the esophagus, the oral cavity, the cervix and the lung. The fluorescence imaging apparatus for an endoscope of the present invention utilizes a selected range of excitation light wavelengths and a fluorescence normalization method. These choices provide for an improved endoscope which requires one, non-intensified, imaging detector at the distal tip of a video endoscope for both white light imaging and fluorescence imaging. The imaging detector can be a pixilated integrated circuit device such as a CMOS imaging device, charge coupled device (CCD) or other small two dimensional imaging sensors that can detect in the visible and infrared ranges.

The system of the present invention has the ability to switch back and forth between white light and fluorescence visualization methods electronically, requiring no moving parts within the endoscope itself. The elimination of cameras and imaging optics at the proximal end of the endoscope significantly improves its handling characteristics. A computer-based imaging system allows quantitative images of the tissue to be displayed, at refresh rates of up to 10 Hz or higher, for example. A preferred embodiment uses easily seen, false-color overlays which indicate the likely areas of tissue dysplasia. By "false color" we mean that a color value is assigned to a particular level of fluorescence intensity for each pixel. A data processing system can be programmed to provide a color system suitable for imaging a given type of tissue condition. The system can be used either with color wheel (e.g., using a monochrome CCD) video endoscopes or with color imaging sensor endoscopes.

In a preferred embodiment, near ultraviolet light is chosen as the excitation wavelength, as described in Wang, et al., US Provisional Application No. 60/072,455, January 26, 1998. This choice in the range of 300-420 nm reduces or eliminates the need for additional filters between the tissue and the imaging detector

due to the fact that standard, electronic imaging sensors, as used in video endoscopes, are insensitive to the excitation light. Visible red light, to which the imaging detector is very sensitive, is chosen to illuminate the tissue for the purpose of acquiring a reference image. This reference light is passed through the same optical guide as the excitation light, and illuminates the tissue with the same normalized spatial distribution and angular distribution as the excitation light. By correlating the spatial intensity and angular distribution of light used for the reflected image and the light for the fluorescence image, a more accurate and diagnostically useful imaging system is provided. This system of the present invention allows the reference image to be used to normalize the fluorescence image so that local reductions in fluorescence intensity can be accurately quantified. In another preferred embodiment, using color wheel (monochrome CCD) video endoscopes, the reference light and excitation light are applied sequentially. In another preferred embodiment, using color CCD video endoscopes, the fluorescence excitation light and reference light are applied simultaneously.

An additional reference image can be taken using a third color channel of the video colonoscope. The reference light directed onto the tissue for this image is predominately blue in color, with a central wavelength and bandwidth which approximates the fluorescence spectrum induced by the UV excitation light. The ratio B/R of this blue reference image, B , to the red reference image, R , on a pixel-by-pixel basis, identifies areas of the imaged tissue where chemical species (predominately hemoglobin) absorb some of the intrinsic fluorescence before it can exit the tissue and be imaged. The acquisition of the ratio F/R of the fluorescence image, F , to the red reference image, R , consequently provides two parameters, F/R and B/R , that can be used on a pixel-by-pixel basis for determining whether abnormal tissue such as dysplasia is present in the imaged region of interest.

BRIEF DESCRIPTION OF THE DRAWINGS

Figure 1 is an illustration of the fluorescence spectrum of normal and dysplastic colon tissue due to excitation with UV light at 370 nm.

Figures 2a and 2b are diagrams of the fiberoptic delivery probe system in accordance with the present invention.

Figure 3 is a schematic diagram of an embodiment of the fluorescence imaging endoscope of the present invention which uses a separate fiberoptic light guide to deliver the UV excitation light through a biopsy channel of a color-wheel (monochrome CCD) video endoscope.

Figure 4 is a flow chart for the production of the output image.

Figure 5 illustrates a timing diagram for the image acquisition, analysis and display process in accordance with the present invention.

Figures 6a - 6c illustrates the UV output from a mercury arc lamp as a function of pulsed current in accordance with the present invention.

Figure 7 is a schematic diagram of a computer-controlled, pulsed lamp power supply in accordance with the present invention.

Figures 8a - 8d illustrate the details of opto-mechanical elements of the pulsed light source in accordance with the present invention.

Figures 9a - 9c illustrate the construction of the optical delivery fiber used in accordance with the present invention.

Figures 10a - 10b are schematic diagrams of a preferred embodiment of the imaging system which uses a separate fiberoptic light guide to deliver the UV excitation light through a biopsy channel of a color CCD video endoscope.

Figure 11a - 11d are schematic diagrams of a preferred embodiment of the imaging system which integrates the UV-transmissive illumination guide into the endoscope and connects an external UV excitation source to a standard white light source with an adapter module.

Figure 12 is a schematic diagram of a preferred embodiment of the imaging system which fully integrates the excitation light source and white light source into

an illumination system for an endoscope fitted with UV-transmissive illumination guide.

Figures 13a and 13b illustrate another preferred embodiment of an illumination and collection endoscope imaging system.

5 The foregoing and other objects, features and advantages of the invention will be apparent from the following more particular description of preferred embodiments of the invention, as illustrated in the accompanying drawings in which like reference characters refer to the same parts throughout the different views. The drawings are not necessarily to scale, emphasis instead being placed upon
10 illustrating the principles of the invention.

DETAILED DESCRIPTION OF THE INVENTION

The present invention relates to systems and methods for imaging autofluorescence from epithelial tissue to highlight regions of dysplasia. The tissue autofluorescence imaging system in accordance with the present invention improves
15 upon the previous state of the art in that, in its simplest form, it can be added on to an existing video endoscope system with no modifications to the endoscope itself, and with only the addition of a shutter in the optical path of the endoscope's visible light source. The handling characteristics of the endoscope are thus not adversely affected by the addition of image intensifiers and external cameras at the proximal
20 end as required by fluorescence imaging systems currently on the market. Switching between normal visible light imaging, in full color, and fluorescence imaging is accomplished by an electronic switch rather than a physical manipulation by the clinician, also required by current systems. The resulting fluorescence video image is processed by a computer so that the diagnostic image seen by the clinician
25 consists of a familiar visible light image (in gray scale) with a false color overlay indicating those areas in the image where the fluorescence from the tissue is reduced in comparison to the fluorescence from normal tissue. This image is much easier to interpret than the combined red/green raw fluorescence images provided by current systems, particularly for red/green color-blind clinicians.

The system of the present invention uses only a single imaging detector at the distal end of the endoscope for acquiring normal color images, fluorescence images and visible reference images. The use of a camera at the distal end is made possible by using fluorescence excitation light at ultraviolet to deep violet wavelengths to which the CCD camera is insensitive or can be made insensitive using a fixed filter. This allows broadband collection of the resulting tissue autofluorescence from blue to red wavelengths, resulting in sufficient light for effective video signals without the need for additional image intensification. Fluorescence imaging in this fashion has been described *in-vivo* by Wang, et al. as described in U.S. Application No. 09/238,664, entitled "Fluorescence Imaging Endoscope" filed on January 26th, 1999, the entire contents of which is incorporated herein by reference.

Previous autofluorescence imaging systems have depended upon imaging very weak fluorescence at red wavelengths to provide an image to compare with the fluorescence image taken at blue-green wavelengths. Additional image intensification is particularly necessary to provide a usable red fluorescence images. The autofluorescence imaging system and method of the present invention avoids this expense and difficulty by supplying additional visible red light illumination to the tissue for the purpose of obtaining a reference image. To be effective, however, the UV excitation light and the visible reference light must be delivered to the tissue through a common optical guide and exit the same illumination aperture with the same angular distribution. This requires a careful design of the excitation and reference light source optics. The processing of the reference image includes other features, such as histogram analysis, to eliminate artifacts such as visible specular reflections which do not occur in fluorescence images.

The combination of the system features in accordance with the present invention can be realized in several different ways, depending upon the endoscope with which the autofluorescence imaging system is combined. Some of these features can also be built into the endoscope itself in stages as the technology is accepted by clinicians.

In a preferred embodiment, the excitation illumination and reference illumination are generated by the same arc lamp source and delivered to the tissue through the same fiberoptic probe passed through a biopsy channel of the endoscope. In other embodiments the excitation light can be generated with a stand-alone source but delivered to the tissue through UV-transmissive illumination bundles built into the endoscopes as replacements to the standard, UV-absorbing glass bundles. The reference illumination is derived from the normal white light illumination source by switching in a red-pass filter that absorbs light at blue and green wavelengths which can interfere with imaging the predominately blue fluorescence. A deep red light source or near infrared (e.g., at 670 nm or above) can be used without affecting the normal visible illumination and can be combined with a dichroic element for delivery along a common optical path. In further embodiments, the excitation light source can be combined with the standard, white light illumination source and delivered to the tissue through an endoscope with UV-transmissive fiberoptic illumination bundles. A rotating wheel light source can also be modified to produce four colors, ultraviolet, blue, green and red, to perform both the fluorescence imaging and the visual imaging.

Figures 2a and 2b show the general components of a preferred embodiment in which the excitation light and the visible reference light are delivered to the tissue through a separate fiberoptic probe. The fiberoptic probe 200 is passed through a biopsy channel of a standard video endoscope 202 with its tip finally placed at or near the distal tip of the endoscope 204. When a footswitch 206 is depressed by the clinician, the normal white light illumination from the endoscope's light source and video processor 208 is switched off with a shutter. This white light normally illuminates the tissue through the two fiberoptic illumination ports 210 and 212 on the distal tip of the endoscope. Simultaneously, a complementary shutter in the excitation and reference illumination source 214 is switched on so that excitation and reference light can pass through the fiberoptic probe 200. The excitation and reference light exits the tip of the fiberoptic probe 216 and illuminates the tissue 218. The video image detection system 220 transmits the resulting fluorescence

image signal and reference image signal back through the endoscope 202 to the video processor 208 where they are converted into different color channels R and B) in a standard R,G,B National Television Standards Committee (NTSC) video format. These two channels are digitized with a video framegrabber in the computer system 222. The digitized fluorescence and reference images are processed together in real time to quantify image regions where the fluorescence is reduced compared to normal tissue. Reduced fluorescence is the primary indicator of dysplasia. Areas of the tissue which are likely to be dysplastic are highlighted with false color in a processed image of the tissue which is displayed on a computer monitor 224 and updated at a rate of up to 10 Hz. The preferred embodiment shown in Figures 3a and 3b is thus an add-on component to an existing endoscope/ video processor system which only requires the addition of an internal shutter to the white light source in the endoscope systems video processor. For color-wheel (monochrome CCD) video endoscopes, the excitation light source 214 is constructed to provide sequential excitation and reference illumination as described in more detail hereinafter. For color-CCD video endoscopes the excitation light source 214 provides simultaneous excitation and reference illumination, also described in more detail hereinafter.

The autofluorescence imaging system demonstrated by Wang, et al. used an argon-ion laser as a UV excitation source. Other laser sources can be used including solid state lasers, such as gallium nitride laser diodes, operating at wavelengths in the range of 380 nm to 420 nm, which have smaller size and low power operation. The systems in accordance with the present invention uses a mercury arc lamp as a source of UV excitation with a spectral band around the 365 nm mercury line. The mercury arc source is smaller, and less expensive than the argon-ion laser, requires relatively little power and is air-cooled. For the autofluorescence imaging systems used with color wheel (monochrome CCD) video endoscopes, the current to the arc lamp can be pulsed. In the color wheel video systems, the normal light source provides red, green and blue light pulses sequentially during a 33 ms video frame which are combined with the video processor to provide a color image. In the autofluorescence imaging mode these systems replace the normal blue light pulse

with a UV pulse and the green light pulse with a (nominally red) reference light pulse transmitted through the same optical fiber as the UV pulse. By pulsing the arc lamp current for the 8 ms period that the blue light is normally on, the UV source can provide as much or more excitation fluence to the tissue than it could if the CW
5 lamp intensity was integrated over the full 33 ms video frame period.

With both color wheel video systems and color CCD video systems the autofluorescence imaging mode is initiated with a footswitch which controls complementary shutters on the excitation/ reference light source and the normal endoscope white light source. Shutting off the normal endoscope illumination is
10 required in order to image the weak fluorescence from the tissue. Two different autofluorescence imaging modes can be used. In one mode, a single frame of the false color overlay from the processed autofluorescence image is combined with the immediately preceding (or following) color image and frozen on the computer screen. In the second mode the processed fluorescence image is updated
15 continuously (at about 7.5 Hz to 10 Hz allowing for the image processing time) for as long as the footswitch is depressed. In the continuous operation mode the visible image is shown as a gray scale (since it is taken with a monochrome reference illumination) with false color overlays showing regions of likely dysplasia.

Figure 3 shows an overall schematic diagram of a preferred embodiment of
20 the fluorescence imaging system for use with color wheel (monochrome CCD) video endoscopes. For normal mode of operation of the color wheel video endoscope, the distal tip of the endoscope 300 is inserted into the hollow organ of the body to observe an area of tissue 302 which may contain a region which is dysplastic 304. The CCD imaging device and lens subsystem (the video camera) 306 is flanked on
25 either side by the normal illumination apertures 308 and 310.

The normal illumination light is carried by a fiberoptic bundle 312 which extends through the length of the endoscope and bifurcates near its distal tip, terminating at apertures 308 and 310. A shutter 314 between the illumination source and the fiberoptic bundle, controlled by a digital signal 315, allows the white light
30 illumination to be turned off without turning off the source lamp 316. In the type of

endoscope shown, the normal color image is obtained by combining three, successive images taken with red, green and blue light pulses provided by a rotating filter wheel 318. The CCD detector in this type of endoscope is sensitive to all wavelengths between 400 nm and 700 nm, but is insensitive to UV excitation wavelengths around 365 nm which are used to excite the autofluorescence. This is due both to the design of the silicon sensor array and to the choice of optical materials used to physically protect the surface of the array. The CCD detector continuously integrates all of the light which falls on its surface so the illumination must be shuttered off while the CCD rows are shifted down to the readout electronics or else a streaking effect will be seen on the image. The red, green and blue light pulses have a duration of about 6 ms followed by a dark period of 5 ms during which the camera pixels are read out, yielding a total video frame period of about 33 ms or 29.97 frames per second to meet the NTSC standard. The analog readout signal from the CCD camera is carried through a cable 320 to the endoscope video processor 322.

The three successive monochrome images are digitized and combined into a standard color video signal at the end of the video frame. The processor has two groups of standard red, green, blue (RGB) plus synchronization outputs. One output group of color signals 324 goes to the endoscopes video monitor 326 to display the normal color image of the tissue 328. Another group of color signals 330 goes to a video framegrabber 332 in the fluorescence imaging computer system 334 which will acquire and process the autofluorescence and reference images. A standard composite color signal output 336 from the video processor goes to a synchronization circuit 338, based on a National LM1881 Video Sync Separator. This synchronization circuit 338 determines when the interlaced even and odd fields occur in the video signal and outputs a binary digital signal 340 which is high during the odd field and low during the even field. This signal 340 is used throughout the fluorescence imaging system to synchronize its functions to the timing set by the endoscope's video processor.

To engage the autofluorescence imaging mode of operation, the clinician depresses footswitch 341 which sends a signal on cable 342 to the computer 334. At the appropriate time, synchronized to the next opportunity as determined by the synchronization signal 340, a signal is sent out on the shutter trigger line 315 to shutter 314 on the normal illumination source and the shutter 343 on the excitation/reference light source. These shutters are complementary, so that the signal on line 315 opens shutter 343 and simultaneously shuts shutter 314. The excitation and reference light pulses, generated by the rotating shutter wheels in the optical group 350, then pass into the fiberoptic probe 344. Probe 344 is inserted into a biopsy channel opening 345 and slid down the channel until its end window 346 is either at or just beyond the distal tip of the endoscope 300. The excitation and reference light pulses illuminate the central portion of the endoscope's visual field. The angular extent of the illumination depends on optical elements 350, the fiber numerical aperture, and the optical characteristics of the fiberoptic probe end window 346.

The excitation and reference light pulses must occur during two of the three normal illumination pulse periods for the autofluorescence and reference images to be properly acquired. These images appear on the next video output frame on two of the three video output channels from the endoscope video processor. The appropriate timing is accomplished by rotating shutters 352 and 354 in the excitation (UV) and reference (RED) light paths, respectively, of the optical group 350. The shutters are driven by direct current (DC) motors 356 and 358 which rotate at a speed which can be controlled by varying their supply voltage. A fiducial hole near the rim of the shutters, combined with an optical source and detector, generates a reference pulse for each which marks its phase as it rotates. A phase-locked loop (PLL) 362 and 360 for each motor, 358 and 356, respectively, adjusts the motors voltage so that the reference pulse for each matches the rising edge of the synchronization pulse 340 marking the beginning of the odd video field.

By positioning the hole in the shutters appropriately the excitation light pulse and the reference light pulse can be timed to the exposure periods of the camera. The excitation pulse is timed to match the normal blue exposure since this exposure

period is somewhat longer (8.1 ms) than the others. The reference light pulse is timed to match the normal green light exposure (5 ms) since it is the next longest of the normal exposure periods. The normal red exposure period, which immediately follows the normal blue exposure period, is not currently used, but can be employed to obtain additional autofluorescence or reference images or an additional visible reflectance image for other spectroscopic analyses. During the excitation exposure the current to the mercury lamp can be boosted to a higher level to increase the excitation light output. The lamp power supply utilizes a DC current section 366 to maintain the idle current and start the lamp.

10 A computer-controlled, pulsed current section 368, connected in parallel with the DC section 366, can rapidly switch in multiple, parallel-coupled, constant-current sources to vary the output power of the lamp as required by the imaging system. The current pulses are synchronized to the video system using the same synchronization pulse 340 that serves to lock the rotating shutters. The computer digital input/output (I/O) section 370 outputs a digital pulse 372 which is combined with the timing pulse 340 to boost the lamp current during the excitation exposure. The number of constant current sections which are triggered in parallel can be varied by a set of digital control lines 374 as required. If the tissue is sufficiently close to the excitation/ reference delivery probe window 346 then no boost may be required. If the peak value of the autofluorescence image drops to a minimum acceptable level, as determined by the computer image analysis program 376 as monitored by the overall control program 378, then additional current boost sections are enabled as required for the next exposure.

25 Once the autofluorescence image and the reference image have been acquired by the digital framegrabber in the computer then the analysis can begin. A reflectance (non-fluorescing) image taken with an endoscope camera system measures the brightness of the tissue surface in its field of view. To the extent that the tissue surface is a Lambertian (non-specular) reflector, the reflectance image signal (as digitized by the video framegrabber for each discrete pixel in the video image) is proportional to distance of the tissue from a single illumination source (or

a weighted distance from multiple sources), and to the integrated energy of the excitation illumination during that video frame. The end window 346 of the excitation/ reference delivery probe is not in the direct line-of-sight from the camera lens to the tissue so there will be visible shadows. A reflectance image can thus be used to measure the excitation illumination at the tissue surface visible to the camera, including the presence of shadows in the autofluorescence image. Note that this is only true if the excitation illumination and the visible illumination emanate from the same aperture with the same transverse intensity profile and angular divergence as provided by the design of the source as described herein above. A visible reflectance image taken with light from the illumination bundles of a standard endoscope, for instance, is not acceptable for determining the excitation illumination from a separate optical fiber passed through a biopsy channel of that endoscope. Note also, however, that the same aperture/ same divergence condition can be satisfied by passing both the excitation and the reference light pulses through the illumination bundle of the endoscope.

The sequence of steps for obtaining a false color indication of dysplasia probability using the visible reference image along with the autofluorescence image proceeds as follows. The two images are first corrected for the gamma factor applied to the video signal by the video processor which is generally connected to a video monitor rather than a framegrabber. This insures that the digitized images acquired by the framegrabber in the computer are linear functions of the illumination fluence (time-integrated intensity). The two images are then normalized to their peaks, which is generally a region of non-dysplastic tissue somewhere in the visual field. There are a few pixels in the reference image which are saturated due to specular reflections of the reference illumination. These are effectively eliminated by producing a histogram of the reference image and normalizing the image to a peak value. This generally includes about 99% of the pixels.

All normalized reference image pixels above a value of 1 (the specular reflections) are then reset to a value of 1. There are no specular reflections in the autofluorescence image so such a histogram-based normalization is not necessary.

On a pixel-by-pixel basis the autofluorescence image value is then divided by the corrected reference image value to produce a ratio image. This division is performed only when the autofluorescence and reference image pixel values are above a minimum threshold condition to insure that the analysis is not attempted with too little illumination to provide a reliable measurement.

If a ratio image pixel value falls below a predetermined value (typically one-half to one-third) then that pixel represents a region of reduced fluorescence on the tissue surface which is indicative of dysplasia. The corresponding pixel in the processed output image can then be set to a false color state to indicate the relative probability of dysplasia. If the ratio image pixel value is below $1/3$ the red value of the corresponding processed output image pixel it is set to the value of the reference image and the green and blue values of that pixel are set to zero (it is a shaded red, indicating a high probability of dysplasia). If the ratio image pixel value is $1/2$ to $1/3$, then the green value of the processed output image pixel it is set to the reference image value and the red and blue values are set to zero (it is a shaded green, indicating a moderate probability of dysplasia). If the ratio image pixel value is above $1/2$ then the red, green and blue values of the processed output image pixel are all set to the reference image value (it is a shaded gray, indicating the probability that the tissue is normal). This processed output image 386 is displayed on an LCD monitor 384 attached to the system computer 334.

Figure 4 illustrates a process sequence 400 after a patient has been prepared the endoscope inserted into the body cavity or human and the distal end thereof has been positioned for imaging of a region of interest. In this particular example, a visible reference image is obtained 402. This reference data is corrected 404, a histogram is generated 406, the data is normalized 408, selected pixels are reset 410 and a threshold value is applied 412. After a fluorescence image is obtained 420, the image is corrected 422, normalized 424, a threshold applied 426, and a ratio image is generated 430. The resulting output image or representation is then compared 432 to a reference, and a given region is determined to be normal 440 or dysplastic 450.

Figure 5 shows a timing diagram for one cycle of the imaging process. The maximum video output rate is 29.97 Hz as set by the NTSC standard. The diagram indicates a system which requires one frame time for acquisition, between two and three frame times for analysis, and a fraction of a frame time for transferring the result to the output image buffer. The resulting update rate for the analysis image is 7.5 frames per second. With a single, fast processor, the analysis time can be reduced and the update rate correspondingly increased to 10 or 15 frames per second. The image analysis can also be performed at output rates up to 30 frames per second by, for example, using two processors operating in parallel with only a few frames of delay between acquisition and display. Increased update rates, however, require a reduced upper limit on the pulsed current to the lamp in order to maintain an average lamp dissipation of no more than 100 W. Figure 5 also shows how the mercury lamp current is boosted only during the blue exposure periods which result in autofluorescence images. At other times the lamp idles at a reduced power.

Figures 6a and 6b show that the UV output power from a 100 W Hg arc lamp, pulsed for 8 ms above a 70% idle current, is essentially a linear function of its input power to at least a factor of 5 over its nominal rated power. Boosting the lamp current during the UV exposure period increases the lamp output power which, in turn, allows a larger area of tissue to be scanned for dysplasia. It also provides a means for adjusting the output of the lamp for an optimum video exposure on a pulse-to-pulse basis. Since the lamp discharge maintains a nearly constant voltage drop across the arc regardless of current, the lamp output power is essentially proportional to current. About 70% power to the lamp must always be maintained, however, to keep the mercury in the vapor phase. Figure 6c shows that pulsing the current to a factor of 5 over the rated CW current can be repeated continuously at a rate of 7.5 Hz.

Figure 7 shows a block diagram of the lamp power supply as it connects to the mercury arc lamp 700. The DC idle current circuit 702 and high voltage (HV) starting pulse circuit 704 are handled by a single commercial power supply designed

for CW 100 W mercury arc lamps. The constant current, triggered pulse circuit 706 is designed to match the specific requirements of the fluorescence imaging endoscope. Four of these circuits are connected in parallel so that the output of the arc lamp can be set digitally to one of 5 different power levels (including idle). Each circuit is composed of a power MOSFET switch which adjusts its resistance to keep the passed current at a fixed level, typically equivalent to the normal DC current of 4 amps. These circuits can be individually triggered by a computer whenever their current boost is required to properly illuminate the tissue. Each circuit draws its current from a storage capacitor 708 charged by a power supply 710 through a current limiter 712. This limited current design minimizes the possibility of fault conditions which could overload the power handling capacity of the arc lamp.

Figures 8a, 8b, 8c and 8d show details of the excitation/ reference illumination source optics. A single mercury arc lamp 800 in Figure 8a is used as the source of both wavelengths because both the UV/violet excitation light and the red/near infrared reference light need to have the same apparent source volume 802 at the distal tip of the endoscope. Small 100 W, mercury arc lamps have arc dimensions of 0.5 to 1.0 mm which are small enough to couple efficiently into the optical fibers used to deliver the light to the tip of the endoscope. Separate lamps can be used for the excitation and reference beams, but the tolerances for the source optics and the alignment of the source arcs are then more critical. In the preferred, single lamp, source design shown in Figure 8a the light collection optic 804 is shown schematically as a single, fused silica, UV-transmitting lens. In practice either a multiple element lens design or a mirror-based Schwarzschild objective is used to reduce optical aberrations on the collected beam. Such collection optics could also collimate light from the focal volume of a commercial mercury arc lamp with a fixed, pre-aligned, elliptical reflector as an alternative source. The collection optics 804 collimate the light from the lamp so that it can be effectively filtered into UV and visible components by means of a dichroic mirror 806 which reflects UV and/or deep violet wavelengths and transmits visible wavelengths.

Separation into two paths is required so that two rotating shutters, 808 and 810, can produce UV excitation and visible reference pulses at different times during the video frame as described above. A UV-reflective surface coating on mirror 812 immediately after the dichroic mirror reflects nearly 100% of the desired UV light in that path and dumps much of the undesired visible light reflected from the dichroic into its substrate. Additional UV filters represented by element 814 may include absorptive filters such as Schott UG-1 glass as well as multilayer dielectric bandpass filters centered on the 365 nm line of mercury. The UV path rejects visible light to a high degree since the efficiency of the 460 nm tissue fluorescence is only about 0.1%. Leakage of visible source light during the UV exposure period reduces the contrast of the autofluorescence image. Some correction for such leakage is possible during the digital image processing but corrections always add a small amount of noise to the result.

The first element in the reference path is a weak lens 816 to correct for the lower refractive index of the fused silica collimating lens 804 at visible wavelengths as compared to the UV. Note that lens 816 would not be necessary if a Schwarzschild objective is used at the position of lens 804 since such a design uses only mirrors and is thus totally achromatic. An attenuator can optionally be moved into the reference path at position 817 to prevent saturation of the reference image at close inspection distances. This attenuator must be uniform across the area that the reference beam passes through in order to maintain the uniform angular intensity distribution required of the reference illumination. This attenuation can be accomplished with variable, crossed polarizers, differentially-sliding, linear attenuation wedges or electromechanically-switched, fixed-value attenuators. The visible filter 818 in the reference beam path is less critical than the UV filter when sequential excitation/ reference beams are used such as in the system shown. The reference beam wavelength is chosen to avoid hemoglobin absorption bands since a significant absorption introduces an error into the analysis which assumes that the reference image is equivalent to a measurement of the excitation illumination intensity. The rotating shutters 808 and 810 follow the filters.

Continuing down the visible path after shutter 808 is turning mirror 820 with a broadband visible reflective coating. A second dichroic mirror 822, identical to dichroic mirror 806, recombines the UV excitation beam and the visible reference beam onto a common path. One additional turning mirror 824 which reflects both UV and visible light directs the two beams towards the focusing optics 826 which couples them into the delivery fiber 828. The turning mirror 824 makes the number of reflections in both the UV and visible paths of the system equal. With equal numbers of reflections, any change in the position of the mercury arc 802 relative to the position of the collection optics 804 results in the same angular deviation for both the excitation and reference beams. Equal deviation angles preserves the overlap of the excitation and reference beams on the tissue. The system as shown in Figures 8a - 8c, in which the combined beams exit the optical train in the same direction as the incoming beam, also makes the direction of the output beams invariant under translations and small rotations of the optical train as a whole.

The aperture stop 830 in Figure 8a insures that the UV excitation and visible reference beams have the same angular convergence on entering the delivery fiber 828. The transverse dimensions of the two beams at the position of the focusing optic 826 will invariably be slightly different due to small errors in the positioning of the optics in the two paths and slight differences in the effective emitting volume of the arc at the two wavelengths. The aperture stop 830 is set so that both beams are slightly clipped at their outside edge, insuring that the maximum angular input to the optical fiber is the same for both. Figure 8b shows that, to a great degree, the input angle to an optical fiber, relative to the fiber axis, is preserved upon each reflection within the fiber. Illumination of a fiber with a collimated beam from a single direction generally results in a cone of emitted light from the opposite end of the fiber which has the same angle relative to the axis. The light is spatially averaged over the exit aperture of the fiber but the angles only slowly spread due to bends in the fiber as it passes down the length of the endoscope. The embodiment shown insures that the normalized angular intensity distribution of the excitation illumination and the reference illumination are closely matched as required by the

autofluorescence normalization method. If, for example, the normalized reference illumination locally exceeds the normalized excitation illumination in a region 831 as shown in Figure 8c, then the analysis falsely indicates reduced autofluorescence in this region. If the normalized reference illumination is greater by a factor of two over the normalized excitation then the analysis provides a false positive indication of dysplasia, where the threshold condition is set to 50%. Similarly, if the normalized reference illumination is a factor of two below the normalized excitation illumination in some region, then the analysis indicates normal tissue even if the autofluorescence was actually reduced by the same factor of two in that region.

Such large errors are unlikely in the center of the illumination field but can occur at the edges of the field where the illumination drops to the noise level. The threshold conditions placed on both the reference and autofluorescence images insure that the analysis is not attempted where it is too easily influenced by noise or edge effects. In general, it is described to maintain any variation in the intensity distribution of the reference light as seen at 860 (normalized) at less than 20% relative to the intensity distribution of the excitation light at any point along the optical path between the combiner and the tissue surface. It is most important to achieve minimal variation in intensity at the tissue surface. Thus light exiting the fiber 862 is within the cone defined by the angle θ and variations in intensity at along the wavefront 864 should be less than 20% to minimize the risk of false image.

Figure 8d shows a more detailed view of the rotating shutter design in the excitation/ reference source. Each blade rotates once in 33.3 ms, the period of one video frame. The angle subtended by the aperture 832 on the excitation shutter blade 810 corresponds to the 8.1 ms period for the normal blue exposure. The timing of the UV exposure relative to the video frame is set by the phase-locked loop matching the rising edge of the optical pulse through the excitation blade fiducial 834 to the rising edge of the odd field mark on the synchronization signal 340 as shown in Figure 3. The apparent position of the excitation beam in the optical train is marked by the dotted circle 836. For the relative dimensions of the beam diameter to shutter diameter shown the total rise time of the light pulse is 1.7 ms which is

acceptable compared to the total length of the pulse. The blade 810 is thin, light in weight and opaque, made from a material such as stainless steel or glass-fiber reinforced epoxy.

Minimizing the moment of inertia simplifies locking the shutter's rotation rate with the PLL. Being opaque is important to prevent light leakage. The cutouts 838 around the central hub 840 of the wheels leave thin blade flexures 842 which allow the wheel to flex slightly as it rotates to stay perpendicular to the rotation axis of the motor, minimizing vibration and stress on the motor bearings. The continuous outer edge of the shutter serves to carry the timing fiducial at the maximum distance from the axis where it is most sensitive and serves a safety function as well in the case of an inadvertent touch during assembly and testing. The aperture in the visible reference shutter 844 is positioned to match the normal green or red period of the video timing sequence. This instrument can either provide a single false color image for each footswitch signal or a series of processed images at a submultiple of the 30 Hz video frame rate.

A standard personal computer system can process the acquired data and update the false color image at 10 frames per second which is fast enough to preserve the impression of smooth motion and slow enough to allow a significant increase in the pulsed lamp current. Operation at lower submultiples of 30 Hz (7.5 or 6 Hz) still provides adequate real-time feedback and permits higher pulsed currents (and thus brighter UV illumination) without exceeding the nominal 100W average power dissipation in the lamp.

Figures 9a - 9c show details of the optical fiber probe which delivers the UV and reference light pulses from the source described above, through a biopsy channel in the endoscope, to the tissue at the distal tip of the endoscope. The delivery fiber 900 must transmit both the UV and the reference light wavelengths, couple efficiently to the light source and be sufficiently flexible so that it doesn't significantly affect the flexibility of the endoscope tip.

The fibers can be made with multiple, small-diameter fused silica fibers, but the preferred system uses a single, UV-transmitting acrylic fiber with a diameter of

from 1.5 to 2 mm. One particular acrylic fiber that is suitable is the Raytela Polymer Optical Fiber made by Toray Industries, Inc. The single fiber system increases coupling efficiency by eliminating the packing fraction losses of multifiber bundles. Eliminating the multiplicity of internal cavities also increases the reliability of the disinfection process between procedures.

A window 902 at the tip of the probe is held on with a sleeve 904 and low-fluorescence epoxy in the gaps 906, 908 and 910 as shown in Figure 9b. A thin, bio-compatible, sheath of heat-shrink tubing protects the thin cladding of the optical fiber. A band of epoxy at position 914 in the gaps between the fiber 900, the sheath 912 and the sleeve 904 seals the probe for disinfection and holds the sheath in place. This type of fiber has a numerical aperture, NA, of 0.5 which means that the light is emitted over a 60 degree, full-angle cone. Probes with a plane window illuminate only about half of the 120 degree maximum field of view for a typical endoscope.

Figure 9c shows a negative lens 916 used in place of the plane window 902 with an air gap 918 between the lens and the end of the optical fiber. The negative lens increases the illuminated field of view at the expense of excitation illumination intensity. Additional negative lenses would increase the field further. Probe designs such as that of Figure 9c are optimized for scanning large areas while the plane window design of Figure 9b is optimized for scanning more localized regions of dysplasia. These windows or lenses can be made of fused silica or UBK-7 to optimize UV transmission. Millimeter thicknesses of common glasses such as BK7, however, do not absorb a significant fraction of the excitation illumination. The lenses or windows can also be made from a UV-transmitting, blue-blocking, red-transmitting glass such as Schott UG-1 if the level of blue fluorescence in a particular type of delivery fiber degrades the fluorescence image. Such additional filtration has not been required in probe designs using the Raytela fiber.

In a preferred embodiment, constructing the above probe from a UV-transmissive plastic will reduce the cost, compared to a fused silica construction, to the point where the entire probe can be disposed after a single use.

Figures 10a and 10b show how the autofluorescence imaging system shown in Figure 3 can be modified for use with a standard, color-CCD video endoscope 1000. In this type of endoscope the image detection system 1002 detects red, green and blue light simultaneously with discretely filtered pixels on the CCD detector.

- 5 The illumination source 1008 for this type of video system emits continuous, broadband, white light which still needs to be turned on and off with a shutter 1010 for use with the modified autofluorescence imaging system. In this embodiment the excitation/ reference light source 1018 produces the two wavelengths simultaneously with the optical train 1020, with a high rejection of light at blue and green
- 10 wavelengths (greater than 1000:1) where the autofluorescence peaks. The spectrum of the continuous illumination is shown in Figure 10b.

- For autofluorescence detection the complementary shutters 1010 and 1024 are triggered as before, causing the tissue to be illuminated with both the excitation light and reference light simultaneously. The UV-induced autofluorescence
- 15 (primarily at 460 nm) is then detected by the blue-responsive pixels in the CCD camera. The red reference reflectance image is simultaneously detected by the red-responsive pixels. Color CCD cameras typically use electronic shuttering so that they don't need a dark period for their readout. In this embodiment, the rotating shutters in the excitation/ reference light source 1018 are either stopped in an open
- 20 position with a detent, or eliminated altogether if the autofluorescence system is to be used only with color video endoscopes. The UV illumination thus lasts for the full 33 ms of the frame, increasing the integrated visible fluorescence and reference signals and eliminating the need for a pulsed current supply. The lamp current supply 1022 could still be pulsed occasionally for single images without overheating
- 25 the lamp. The analysis would proceed as before with the reference image appearing on the red channel of the RGB NTSC signal and the autofluorescence image appearing on the blue channel.

- In the fluorescence image endoscope in accordance with the present invention, both the UV-excitation light pulse and the visible-reference light pulse
- 30 are delivered to the tissue through an optical probe inserted through a biopsy channel

of a standard endoscope. The excitation and reference light could alternatively be passed through the illumination bundle of an endoscope if that bundle were to be modified to transmit UV as well as visible wavelengths. The requirement would remain that the excitation light and the reference light, even if generated by different sources, have the same angular distribution as they exit the distal tip of the endoscope.

Figures 11a - 11d show an embodiment of the autofluorescence imaging system which uses a UV-transmissive endoscope, a standard white light source for normal imaging, a separate excitation light source and a coupling box. This system is used if a modified endoscope were to be developed before a dedicated light source combining the white light, excitation and reference functions in one unit. In this intermediate instrument, the adapter box 1100 would fit between the standard white light illuminator 1102 and the electrical/ optical connection plug 1104 of the UV-transmissive video endoscope. The optical connection plug 1106 normally fits directly into the visible light source 1102 to collect the illumination light and pass it into the illumination bundle of the endoscope. In this embodiment the optical connection plug 1106 would fit instead into the adapter 1100 and an identical optical connection plug 1108 on the opposite end of the adapter would fit into the visible light source 1102. A set of imaging optics 1110 would transfer the light emitted from the exit aperture of plug 1108 to the entrance aperture of plug 1106. This set of imaging optics would also transfer excitation light from a separate light source, exiting the optical fiber 1112 to the entrance aperture of plug 1106.

The imaging optics set is detailed in Figure 11b. The combination of the excitation illumination onto a common axis with the visible illumination is accomplished with a dichroic beamsplitter cube 1114. An achromat 1118 collimates visible light expanding from a point 1118 on the exit aperture of plug 1108. A second achromat 1120 refocuses this light onto point 1122 on the entrance aperture of plug 1106. A set of fused silica lenses 1124 and 1126 nominally collimate the light from point 1128 on the exit aperture of the excitation delivery fiber 1112. The position of these lenses is actually adjusted to provide the best imaging of point

1128 onto point 1122 since the achromat 1120 will not be corrected for UV wavelengths.

In this embodiment the reference illumination is derived from the normal red illumination light from the visible light source. Filter 1130 is made of red absorbing
5 glass which strongly attenuates wavelengths UV, blue and green wavelengths. Filter 1130 is mounted on a sliding base 1134, shown in Figure 11c, and is switched into the optical path electromechanically when autofluorescence imaging is to be used. The shutter in the UV light source is opened at the same time. The tissue will then be illuminated by both the UV and the reference light either simultaneously if a
10 color video endoscope is being used or sequentially if a color wheel video endoscope is being used. Filter 1132 is clear glass which transmits all of the normal visible wavelengths. Filter 1132 is switched in electromechanically when normal illumination is to be used so that the focal point of lens 1116 remains constant.

For the endoscope illumination bundle 1136 to be both UV-transmissive,
15 flexible and long-lasting it must be made of fused silica optical fibers rather than regular glass fibers. The primary optical difference between the two materials, other than transparency, is that fused silica fibers generally have a lower numerical aperture than glass fibers. This means that light is collected and emitted at smaller angles from the axis in fused silica fibers. The illumination bundle bifurcates at
20 point 1138 and exits the optical ports 1140 and 1142 at the distal tip of the endoscope. The lens elements in the ports 1140 and 1142 must also be modified (made more negative) to maintain the same illumination angle for existing, glass fiber based endoscopes.

The video imaging detection system 1144 at the distal tip of the endoscope
25 transmits its signals back down the electrical wires 1146 within the endoscope to the electrical connector 1147 on the connector plug 1104. A mating connector on the adapter box 1100 collects these signals which are transferred across the adapter 1100 to a connector plug 1148 identical to plug 1147 which completes the electrical connections into the video processor. Analysis of the autofluorescence and reference
30 images then proceeds as before.

Once the UV-capable endoscope is available, the excitation source can be built into the endoscope light source and video processor so that the entire autofluorescence imaging capability is included in the endoscope system itself. This preferred embodiment is shown schematically in Figure 12. The electrical/ optical
5 connection plug 1200 is mated directly to the excitation/ reference/ white light source and video processor 1202. This system operates exactly as the embodiment shown in Figures 11a - 11d, using a dichroic cube beam combiner system 1204 and a reference light filter system 1206, but without the necessity of an external adapter.

The overall reduction in visible blue-green autofluorescence is the primary
10 indicator of tissue dysplasia for the imaging systems described to this point. These imaging systems characterize the autofluorescence reduction with a single parameter for every pixel in the tissue image. The specific causes of this reduction have been described by Zonios et al. in "Morphological Model of Human Colon Tissue Fluorescence," IEEE Trans. Biomed. Eng., 43(2), 113-122, 1996, incorporated
15 herein by reference. The primary fluorophore is collagen, a protein in connective tissue, which is found within the thin mucosal layer and which is the dominant component of the submucosal layer. The autofluorescence in dysplastic tissue is reduced when cells in the mucosal layer enlarge and crowd out the collagen and when hemoglobin is added in the tissue due to increased vascularization. The
20 hemoglobin absorbs some of the autofluorescence. The ability to characterize both the intrinsic fluorescence of the tissue (in the absence of absorption) and the level of hemoglobin in the tissue improves the ability of the imaging system to correctly determine the likelihood of dysplasia.

The imaging systems described above utilizes two of the three available
25 imaging periods in the 33 ms acquisition period of a standard, color-wheel (monochrome CCD) video colonoscope. During one period, the ultraviolet excitation light illuminates the tissue and an autofluorescence image, F, is acquired. During a second, period visible red light illuminates the tissue and a reference image, R, is acquired which determines the intensity distribution of the ultraviolet
30 illumination on the tissue. The ratio of these two images produces a single

parameter, F/R , for every pixel in the image which represents the relative level of overall autofluorescence at that point. The algorithm for highlighting portions of the image where dysplasia is likely is primarily one of assigning appropriate threshold values to the F/R parameter for each pixel.

5 The improved imaging system described below and shown in Figures 13a and 13b utilizes the third imaging period to acquire an additional visible reflectance image to allow for a more complete spectroscopic analysis. In this system predominately blue light is directed onto the tissue with a central wavelength and bandwidth which approximates the fluorescence spectrum induced by the UV
10 excitation light. As before, this blue light exits the same illumination ports and has the same angular intensity distribution as the ultraviolet excitation illumination and the visible red reference illumination. The resulting blue reflectance image, B , indicates regions where chemical species in the tissue absorb light at these blue wavelengths and thus areas of tissue where the intrinsic autofluorescence is reduced
15 by absorption processes. As before, the red reference image, R , is used as the measure of the illumination intensity of all three wavelengths at the tissue surface. The specific parameter to indicate (primarily hemoglobin) absorption on a pixel-by-pixel basis is thus the ratio B/R . The two parameters, F/R and B/R , are available on a pixel-by-pixel basis for determining that the imaged area of tissue indicates
20 dysplasia. Two basic methods can be used. In one, the F/R ratio can be corrected by the B/R parameter to more closely indicate the true value of the intrinsic fluorescence which produces a single, modified F/R parameter to which a threshold is applied as before. A second procedure uses a two-dimensional surface, defined by the F/R and B/R parameters, which can indicate the probability of dysplasia. In both
25 cases the threshold values for determining which are to be highlighted is determined by comparing measured values of F/R and B/R during clinical tests with the results of pathology reports on biopsy samples taken from those sites.

 The system in Figures 13a and 13b uses a mercury arc lamp source 1300 as before. The same optical elements 1302 and 1304 are used to collect and collimate
30 the light produced which includes both ultraviolet wavelengths and the full range of

visible wavelengths. The optical path of the ultraviolet illumination remains the same. A dichroic mirror 1306 reflects most of the UV light and some visible light to a UV-reflective turning mirror 1308. Again, most of the UV light is reflected while most of the visible light is absorbed into the substrate of 1308. A UV bandpass filter 1310 removes almost all of the remaining visible light so that only UV light reaches the timing wheel 1312. The timing wheel 1312 insures that the UV light which continues on is only passed during one exposure period for the video colonoscope. Another dichroic mirror 1314 and a total reflecting mirror (both UV and visible) 1318 direct the UV beam to the focusing lens 1318. An iris 1320 defines the angular extent of the UV light entering the delivery fiber 1322. The delivery fiber 1322 carries all of the illumination wavelengths to the end of the colonoscope to illuminate the tissue.

The difference between the improved autofluorescence imaging system and the systems described previously lies in the path that the visible wavelengths take through the system. Both blue and red wavelengths approach the visible timing wheel 1324. This wheel has two apertures which allow light to pass during the two remaining exposure periods for the video colonoscope. One of the wheel apertures is covered by a red filter 1326 which passes wavelengths appropriate for the reference image, R. These wavelengths are chosen to avoid both hemoglobin absorption bands in the tissue and absorption bands in the optical fiber used to deliver the light to the tissue. The light to be filtered is well collimated so that multilayer dichroic coatings can be used, but simple absorption filters made of colored glass or plastic are adequate. The second aperture on timing wheel 1324 is covered by a broadband blue filter which approximates the spectrum of the intrinsic fluorescence produced by the excitation light. The specific bandpass spectrum of this filter can contain wavelengths absorbed by hemoglobin and be sufficiently reproducible so that the resulting analysis program can be used by other instruments. The second aperture matches the third exposure period of the video colonoscope. Photoelectric detectors responding to light passing through timing holes 1330 and 1332 on the two wheels as shown in Figure 13b provide electrical pulses for the phase-locked loops which

keep the timing wheels synchronized with the video acquisition system of the video colonoscope.

The remaining differences between the improved, three-color diagnostic system and the two-color system include a broadband, reflective mirror 1334 that must
5 reflects both blue and red visible wavelengths and an achromatic, long-focus lens 1336 which corrects for the longer focal length of the fused silica collection lens 1306 at visible wavelengths.

While this invention has been particularly shown and described with references to preferred embodiments thereof, it will be understood by those skilled
10 in the art that various changes in form and details may be made therein without departing from the scope of the invention encompassed by the appended claims.

CLAIMS

What is claimed is:

1. A fluorescence imaging system comprising:
 - a light source for producing excitation light that induces visible
 - 5 fluorescence in tissue and a reference light;
 - an optical combiner that combines said excitation light and said reference light onto a common optical path, said combined light being coupled into an optical guide that delivers the combined light to the tissue;
 - an image sensor that detects a fluorescence image and a reference
 - 10 image of the tissue; and
 - a data processor that processes the fluorescence image and said reference image to produce a processed output image of the tissue.
2. The system of Claim 1 wherein the light source is an arc lamp.
3. The system of Claim 2 wherein the arc lamp current source is a pulsed
- 15 source.
4. The system of Claim 1 wherein the optical guide is a removable fiberoptic extending through a biopsy channel of an endoscope.
5. The method of Claim 1 wherein the image sensor is located at a distal end of a endoscope.
- 20 6. The system of Claim 1 wherein the excitation light and the reference light are emitted sequentially such that a monochromatic image sensors detects a fluorescent image during a first time period and detects a reflected image during a second time period.

7. The system of Claim 1 wherein the excitation light and the reference light are emitted simultaneously such that respective images are detected by a color-sensitive image sensor, a blue channel detecting the fluorescence image and a red channel detecting the reference image.
- 5 8. The system of Claim 1 wherein the excitation light is in the range of 300 to 420 nm.
9. The system of Claim 1 wherein the light source further comprises a reference light source having a wavelength in a red or infrared range.
- 10 10. The system of Claim 1 wherein the optical guide comprises an optical fiber with a distally mounted lens.
11. The system of Claim 1 wherein the excitation light has an angular orientation that is the same as an angular orientation as the reference light.
12. The system of Claim 1 further comprising a reference illumination of a selected wavelength to provide a second reference image.
- 15 13. The system of Claim 12 further comprising a light source and a first filter wheel that filters light from the light source to provide a first reference illumination wavelength and a second illumination wavelength.
14. The system of Claim 1 further comprising a filter wheel having a filter that filters light from the light source to provide fluorescence illumination.

15. A method for imaging tissue fluorescence comprising:
detecting a fluorescence image of tissue and a reference image of the
tissue; and
processing said fluorescence image together with said reference image to
5 produce an output image of the tissue.
16. The method of Claim 15 further comprising providing an arc lamp light source.
17. The method of Claim 16 further comprising pulsing the arc lamp current source.
18. The method of Claim 15 further comprising sequentially the images with a
monochromatic image sensor.
- 10 19. The method of Claim 15 further comprising detecting images with a color-
sensitive image sensor, the sensor having a blue channel detecting an
autofluorescence image and a red channel detecting the reference image.
20. The method of Claim 15 further comprising coupling excitation light and a
reference light to an optical fiber such that a variation in a normalized intensity
15 of the reference light and a normalized intensity of the excitation light is less
than 20% at any point in a wavefront along the optical path between a combiner
that combines the excitation light and the reference light.
21. A method for imaging tissue fluorescence comprising:
providing excitation light having a wavelength in a range of 300 nm to
20 420 nm;
providing a reference light;

combining said excitation light and said reference light onto a common optical path such that an intensity of the excitation light varies less than 20% relative to a normalized intensity of the reference light at any point along the optical path;

5 detecting a fluorescence image of the tissue due to the said excitation light and a reference image of the tissue due to reflected reference light with an imaging sensor at a distal end of an endoscopic probe; and

 processing said fluorescence image and said reference image to produce an output image of the tissue.

- 10 22. The method of Claim 21 further comprising determining a ratio of the fluorescence image and the reference image to provide a processed image.
23. The method of Claim 21 further comprising adjusting the relative intensity or angular distribution of the reference light relative to the excitation light.
24. The method of Claim 21 further comprising detecting first and second reference
15 images having different wavelengths.
25. The method of Claim 24 further comprising detecting blue and red reference images and forming a third reference image from a pixel-by-pixel ratio of the blue and red images.

1/13




Figure 1

2/13




Figure 2a

Figure 2b

SUBSTITUTE SHEET (RULE 26)

3/13




Figure 3

SUBSTITUTE SHEET (RULE 26)




Figure 4

SUBSTITUTE SHEET (RULE 26)




Figure 5

6/13




Figure 6a




Figure 6b




Figure 6c

SUBSTITUTE SHEET (RULE 26)

7/13





Figure 7






Figure 9a




Figure 9b




Figure 9c

10/13




Figure 10b

Figure 10a

SUBSTITUTE SHEET (RULE 26)

11/13




Figure 11c




Figure 11b




Figure 11a




Figure 11d

SUBSTITUTE SHEET (RULE 26)

12/13




Figure 12

13/13




Figure 13a




Figure 13b

SUBSTITUTE SHEET (RULE 26)

INTERNATIONAL SEARCH REPORT

Intern. Appl. No.

PCT/US 00/01673

A. CLASSIFICATION OF SUBJECT MATTER

IPC 7 A61B5/00

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

IPC 7 A61B 601N

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X A	US 5 187 572 A (NAKAMURA KAZUNARI ET AL) 16 February 1993 (1993-02-16) column 15, line 37 -column 16, line 3 column 19, line 12 - line 31 column 3, line 65 -column 4, line 39 column 9, line 22 - line 28	15-18, 20,21 1-6, 8-11,14, 23
X A	US 5 769 792 A (HUNG JACLYN Y-C ET AL) 23 June 1998 (1998-06-23) column 4, line 24 - line 35 column 5, line 12 - line 63 column 7, line 1 - line 15 column 7, line 31 -column 8, line 14 -/-	15,18, 20,21,24 1,4,6-8, 11-14, 19,22, 24,25

☒ Further documents are listed in the continuation of box C.☒ Patent family members are listed in annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier document but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

"&" document member of the same patent family

Date of the actual completion of the international search

8 June 2000

Date of mailing of the international search report

16/06/2000

Name and mailing address of the ISA

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040, Tx. 31 651 epo nl,
Fax: (+31-70) 340-3016

Authorized officer

Knüpling, M

INTERNATIONAL SEARCH REPORT

International Application No

PCT/US 00/01673

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 5 413 108 A (ALFANO ROBERT R) 9 May 1995 (1995-05-09)	15
Y	column 4, line 15 -column 5, line 4	1,4,6,8, 11,14
A		2,16,18
Y	DE 196 27 369 A (HEINRICH HERRMANN DR; SPECK UWE DIPL ING (DE)) 9 October 1997 (1997-10-09) column 2, line 30 -column 3, line 61	1,4,6,8, 11,14
A	US 4 449 535 A (RENAULT GUY) 22 May 1984 (1984-05-22) column 2, line 49 - line 63 column 4, line 66 -column 5, line 23	1,4, 7-11,15, 21,22

INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No

PCT/US 00/01673

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 5187572 A	16-02-1993	JP 5084218 A DE 4136034 A	06-04-1993 07-05-1992
US 5769792 A	23-06-1998	US 5507287 A AT 140600 T CA 2042075 A DE 69212382 D EP 0512965 A JP 6054792 A	16-04-1996 15-08-1996 09-11-1992 29-08-1996 11-11-1992 01-03-1994
US 5413108 A	09-05-1995	NONE	
DE 19627369 A	09-10-1997	NONE	
US 4449535 A	22-05-1984	FR 2502783 A FR 2521727 A DE 3210593 A JP 1588724 C JP 2007653 B JP 57188242 A	01-10-1982 19-08-1983 28-10-1982 19-11-1990 20-02-1990 19-11-1982

Form PCT/ISA/210 (patent family annex) (July 1992)



INTERNATIONAL APPLICATION PUBLISHED UNDER THE PATENT COOPERATION TREATY (PCT)

<p>(51) International Patent Classification ⁷ : A61B 5/00</p>	A1	<p>(11) International Publication Number: WO 00/42910</p> <p>(43) International Publication Date: 27 July 2000 (27.07.00)</p>														
<table style="width: 100%; border: none;"> <tr> <td style="width: 50%; vertical-align: top; padding: 5px;"> <p>(21) International Application Number: PCT/US00/01673</p> <p>(22) International Filing Date: 26 January 2000 (26.01.00)</p> <p>(30) Priority Data:</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">09/238,664</td> <td style="width: 30%;">26 January 1999 (26.01.99)</td> <td style="width: 40%;">US</td> </tr> <tr> <td>09/362,806</td> <td>28 July 1999 (28.07.99)</td> <td>US</td> </tr> </table> <p>(63) Related by Continuation (CON) or Continuation-in-Part (CIP) to Earlier Application</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">US</td> <td style="width: 30%;">09/362,806 (CIP)</td> <td style="width: 40%;"></td> </tr> <tr> <td>Filed on</td> <td>28 July 1999 (28.07.99)</td> <td></td> </tr> </table> <p>(71) Applicant (for all designated States except US): NEWTON LABORATORIES, INC. [US/US]; Suite 4850, 800 West Cummings Park, Woburn, MA 01801 (US).</p> <p>(72) Inventor; and</p> <p>(75) Inventor/Applicant (for US only): FULGHUM, Stephen, F., Jr. [US/US]; 9 Driftwood Road, Marblehead, MA 01945-1250 (US).</p> <p>(74) Agents: HOOVER, Thomas, O. et al.; Hamilton, Brook, Smith & Reynolds, P.C., Two Militia Drive, Lexington, MA 02421 (US).</p> </td> <td style="width: 50%; vertical-align: top; padding: 5px;"> <p>(81) Designated States: AE, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CR, CU, CZ, DE, DK, DM, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).</p> <p>Published</p> <p><i>With international search report.</i></p> <p><i>Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.</i></p> </td> </tr> </table> <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p>(54) Title: AUTOFLUORESCENCE IMAGING SYSTEM FOR ENDOSCOPY</p> <p>(57) Abstract</p> <p>A system and method for imaging tissue autofluorescence through a video endoscope is described, comprising a light source capable of providing both ultraviolet light capable of inducing tissue autofluorescence and visible light which induces little or no autofluorescence, an optical system to deliver both wavelength bands to the tissue with the same apparent spatial and angular intensity distribution, a means for digitally acquiring the resulting, visible fluorescence and visible reflectance images using a single imaging detector at the distal tip of the endoscope and a means for digitally processing said images to generate a final, false-color image for display which indicates regions of tissue dysplasia. This system can either be added on to an existing video endoscope or integrated into its structure. The combined system can be electronically switched between normal white light imaging and fluorescence imaging.</p> </div> <div style="text-align: center; margin-top: 20px;"> </div>			<p>(21) International Application Number: PCT/US00/01673</p> <p>(22) International Filing Date: 26 January 2000 (26.01.00)</p> <p>(30) Priority Data:</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">09/238,664</td> <td style="width: 30%;">26 January 1999 (26.01.99)</td> <td style="width: 40%;">US</td> </tr> <tr> <td>09/362,806</td> <td>28 July 1999 (28.07.99)</td> <td>US</td> </tr> </table> <p>(63) Related by Continuation (CON) or Continuation-in-Part (CIP) to Earlier Application</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">US</td> <td style="width: 30%;">09/362,806 (CIP)</td> <td style="width: 40%;"></td> </tr> <tr> <td>Filed on</td> <td>28 July 1999 (28.07.99)</td> <td></td> </tr> </table> <p>(71) Applicant (for all designated States except US): NEWTON LABORATORIES, INC. [US/US]; Suite 4850, 800 West Cummings Park, Woburn, MA 01801 (US).</p> <p>(72) Inventor; and</p> <p>(75) Inventor/Applicant (for US only): FULGHUM, Stephen, F., Jr. [US/US]; 9 Driftwood Road, Marblehead, MA 01945-1250 (US).</p> <p>(74) Agents: HOOVER, Thomas, O. et al.; Hamilton, Brook, Smith & Reynolds, P.C., Two Militia Drive, Lexington, MA 02421 (US).</p>	09/238,664	26 January 1999 (26.01.99)	US	09/362,806	28 July 1999 (28.07.99)	US	US	09/362,806 (CIP)		Filed on	28 July 1999 (28.07.99)		<p>(81) Designated States: AE, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CR, CU, CZ, DE, DK, DM, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).</p> <p>Published</p> <p><i>With international search report.</i></p> <p><i>Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.</i></p>
<p>(21) International Application Number: PCT/US00/01673</p> <p>(22) International Filing Date: 26 January 2000 (26.01.00)</p> <p>(30) Priority Data:</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">09/238,664</td> <td style="width: 30%;">26 January 1999 (26.01.99)</td> <td style="width: 40%;">US</td> </tr> <tr> <td>09/362,806</td> <td>28 July 1999 (28.07.99)</td> <td>US</td> </tr> </table> <p>(63) Related by Continuation (CON) or Continuation-in-Part (CIP) to Earlier Application</p> <table style="width: 100%; border: none;"> <tr> <td style="width: 30%;">US</td> <td style="width: 30%;">09/362,806 (CIP)</td> <td style="width: 40%;"></td> </tr> <tr> <td>Filed on</td> <td>28 July 1999 (28.07.99)</td> <td></td> </tr> </table> <p>(71) Applicant (for all designated States except US): NEWTON LABORATORIES, INC. [US/US]; Suite 4850, 800 West Cummings Park, Woburn, MA 01801 (US).</p> <p>(72) Inventor; and</p> <p>(75) Inventor/Applicant (for US only): FULGHUM, Stephen, F., Jr. [US/US]; 9 Driftwood Road, Marblehead, MA 01945-1250 (US).</p> <p>(74) Agents: HOOVER, Thomas, O. et al.; Hamilton, Brook, Smith & Reynolds, P.C., Two Militia Drive, Lexington, MA 02421 (US).</p>	09/238,664	26 January 1999 (26.01.99)	US	09/362,806	28 July 1999 (28.07.99)	US	US	09/362,806 (CIP)		Filed on	28 July 1999 (28.07.99)		<p>(81) Designated States: AE, AL, AM, AT, AU, AZ, BA, BB, BG, BR, BY, CA, CH, CN, CR, CU, CZ, DE, DK, DM, EE, ES, FI, GB, GD, GE, GH, GM, HR, HU, ID, IL, IN, IS, JP, KE, KG, KP, KR, KZ, LC, LK, LR, LS, LT, LU, LV, MA, MD, MG, MK, MN, MW, MX, NO, NZ, PL, PT, RO, RU, SD, SE, SG, SI, SK, SL, TJ, TM, TR, TT, TZ, UA, UG, US, UZ, VN, YU, ZA, ZW, ARIPO patent (GH, GM, KE, LS, MW, SD, SL, SZ, TZ, UG, ZW), Eurasian patent (AM, AZ, BY, KG, KZ, MD, RU, TJ, TM), European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE), OAPI patent (BF, BJ, CF, CG, CI, CM, GA, GN, GW, ML, MR, NE, SN, TD, TG).</p> <p>Published</p> <p><i>With international search report.</i></p> <p><i>Before the expiration of the time limit for amending the claims and to be republished in the event of the receipt of amendments.</i></p>			
09/238,664	26 January 1999 (26.01.99)	US														
09/362,806	28 July 1999 (28.07.99)	US														
US	09/362,806 (CIP)															
Filed on	28 July 1999 (28.07.99)															

FOR THE PURPOSES OF INFORMATION ONLY

Codes used to identify States party to the PCT on the front pages of pamphlets publishing international applications under the PCT.

AL	Albania	ES	Spain	LS	Lesotho	SI	Slovenia
AM	Armenia	FI	Finland	LT	Lithuania	SK	Slovakia
AT	Austria	FR	France	LU	Luxembourg	SN	Senegal
AU	Australia	GA	Gabon	LV	Latvia	SZ	Swaziland
AZ	Azerbaijan	GB	United Kingdom	MC	Monaco	TD	Chad
BA	Bosnia and Herzegovina	GE	Georgia	MD	Republic of Moldova	TG	Togo
BB	Barbados	GH	Ghana	MG	Madagascar	TJ	Tajikistan
BE	Belgium	GN	Guinea	MK	The former Yugoslav Republic of Macedonia	TM	Turkmenistan
BF	Burkina Faso	GR	Greece			TR	Turkey
BG	Bulgaria	HU	Hungary	ML	Mali	TT	Trinidad and Tobago
BJ	Benin	IE	Ireland	MN	Mongolia	UA	Ukraine
BR	Brazil	IL	Israel	MR	Mauritania	UG	Uganda
BY	Belarus	IS	Iceland	MW	Malawi	US	United States of America
CA	Canada	IT	Italy	MX	Mexico	UZ	Uzbekistan
CF	Central African Republic	JP	Japan	NE	Niger	VN	Viet Nam
CG	Congo	KE	Kenya	NL	Netherlands	YU	Yugoslavia
CH	Switzerland	KG	Kyrgyzstan	NO	Norway	ZW	Zimbabwe
CI	Côte d'Ivoire	KP	Democratic People's Republic of Korea	NZ	New Zealand		
CM	Cameroon			PL	Poland		
CN	China	KR	Republic of Korea	PT	Portugal		
CU	Cuba	KZ	Kazakstan	RO	Romania		
CZ	Czech Republic	LC	Saint Lucia	RU	Russian Federation		
DE	Germany	LI	Liechtenstein	SD	Sudan		
DK	Denmark	LK	Sri Lanka	SE	Sweden		
EE	Estonia	LR	Liberia	SG	Singapore		

AUTOFLUORESCENCE IMAGING SYSTEM FOR ENDOSCOPY

GOVERNMENT SUPPORT

This invention was supported in whole or in part by a grant from the National Institute of Health, Grant No. R44CA72626, "Imaging Spectrofluorimeter
5 for Colonic Dysplasia". The Government has certain rights in the invention.

RELATED APPLICATION

This is a continuation-in-part application of U.S. Application No. 09/362,806, filed July 28, 1999 which is a continuation-in-part application of U.S. Application No. 09/238,664 filed on January 26, 1999, the entire contents of the
10 above applications being is incorporated herein by reference.

BACKGROUND OF THE INVENTION

Advanced stage cancers are difficult to treat effectively, so it is important to detect them in their premalignant stages, such as dysplasia or carcinoma in situ. Currently, the most widely used method for early detection uses visual inspection
15 through endoscopes, which relies upon the recognition of the gross architectural changes associated with dysplasia. Visual inspection is less effective in detecting the superficial lesions of flat dysplasia, such as ulcerative colitis and Barrett's esophagus. In these cases, surveillance requires the selection of representative sites for biopsy and subsequent histological analysis. Only a very small fraction of a large
20 surface, such as the colon, can be surveyed in this manner and small areas of dysplasia may go undetected. More efficient methods for detecting flat dysplasia would provide a significant means of reducing cancer morbidity, mortality, and costs.

A promising technique for detecting dysplasia during endoscopy involves illuminating the tissue with light of an appropriate wavelength and observing the resulting fluorescence. Tissue fluorescence occurs at longer wavelengths than the excitation illumination and is typically much weaker, so that spectroscopic techniques are generally required for its detection. Diagnostic methods which use fluorescence information can generally be divided into two groups. The first group of methods observes fluorescence from medications administered to the patient which have accumulated in tumor tissues. The second group of methods observes endogenous fluorescence or autofluorescence arising from substances natural to the tissue itself which change their relative concentrations when the tissue becomes dysplastic. Of the two general fluorescence methods, the first method, requiring a prior application of a drug, is the more invasive. The application of these drugs takes additional time and has the potential of causing adverse side effects. Methods based on autofluorescence detection are less invasive and more suited to endoscopy for screening purposes.

Normal colon tissue, for example, when illuminated with ultraviolet light at 370 nm, exhibits a broad, blue fluorescence with a peak at 450 nm, as shown in Figure 1. This fluorescence is due to collagen, the primary protein of connective tissue, which is found within the thin mucosal layer and which is the dominant component of the submucosal layer. The fluorescence of dysplastic colon tissue, due to changes in its structure and chemistry, is typically $\frac{1}{2}$ to $\frac{1}{3}$ as intense given the same illumination. This reduction in visible blue-green autofluorescence, produced by ultraviolet to violet excitation light, has been identified as a primary indicator of dysplastic tissue. An increase in the relative fluorescence at 680 nm compared to 600 nm is a secondary indicator of dysplasia.

Diagnostic instruments which detect autofluorescence can be divided into two general groups. The first subdivision includes instruments which utilize fiberoptic probes to perform essentially point measurements on the tissue. The second subdivision includes instruments which produce detailed, two dimensional images. Point detection instruments have the advantage of providing more complete

spectral information on the tissue, but are too slow for routine screening of large tissue areas and may miss small regions of dysplasia. Fluorescence imaging endoscopes are more appropriate for screening large tissue areas such as the colon.

Fluorescence imaging systems designed to sense concentrations of
5 fluorescent markers that have been applied to the region of interest are optimized for the measurement of relatively high fluorescence levels, do not describe the additional instrumental features required to measure inherently weak autofluorescence. In particular, they do not describe a method for providing sufficient out-of-band filtering for the excitation illumination that would allow the
10 effective measurement of autofluorescence, which typically is reduced in intensity, compared to the excitation intensity, by a factor of 1000 or more.

Fluorescence imaging endoscopes which are specifically designed to measure autofluorescence can be further divided into groups based on the excitation wavelength chosen and the method for quantifying the degree to which the
15 autofluorescence is reduced. These design choices have a direct bearing on commercial considerations for the instrument since they affect, for example, the number of imaging devices required, the opto-mechanical complexity of the instrument, and the handling characteristics of the instrument in actual use.

Existing fluorescence imaging endoscope systems use visible blue light near
20 440 nm for the excitation wavelength, resulting in fluorescence peaks around 500 nm. These instruments use Helium-Cadmium lasers at 442 nm as a bright, easily controlled source for the excitation light. The high cost of Helium-Cadmium lasers makes them impractical as sources for commercial instruments.

Multiple cameras and mechanically switched optical components and/or
25 filters require the use of endoscopes based on coherent imaging fiber bundles, so that the cameras and filters can be located at the proximal end of the endoscope where there is room for them. The coherent fiber imaging bundles introduce significant light losses and do not provide as sharp an image as those now available with video endoscopes.

SUMMARY OF THE INVENTION

The present invention relates to imaging endoscopes and in particular to an endoscope system and method for imaging autofluorescence from epithelial tissue to highlight regions of dysplasia. The system for detecting dysplastic tissue uses the autofluorescence of mucosal tissue such as that found in the colon, the esophagus, the oral cavity, the cervix and the lung. The fluorescence imaging apparatus for an endoscope of the present invention utilizes a selected range of excitation light wavelengths and a fluorescence normalization method. These choices provide for an improved endoscope which requires one, non-intensified, imaging detector at the distal tip of a video endoscope for both white light imaging and fluorescence imaging. The imaging detector can be a pixilated integrated circuit device such as a CMOS imaging device, charge coupled device (CCD) or other small two dimensional imaging sensors that can detect in the visible and infrared ranges.

The system of the present invention has the ability to switch back and forth between white light and fluorescence visualization methods electronically, requiring no moving parts within the endoscope itself. The elimination of cameras and imaging optics at the proximal end of the endoscope significantly improves its handling characteristics. A computer-based imaging system allows quantitative images of the tissue to be displayed, at refresh rates of up to 10 Hz or higher, for example. A preferred embodiment uses easily seen, false-color overlays which indicate the likely areas of tissue dysplasia. By "false color" we mean that a color value is assigned to a particular level of fluorescence intensity for each pixel. A data processing system can be programmed to provide a color system suitable for imaging a given type of tissue condition. The system can be used either with color wheel (e.g., using a monochrome CCD) video endoscopes or with color imaging sensor endoscopes.

In a preferred embodiment, near ultraviolet light is chosen as the excitation wavelength, as described in Wang, et al., US Provisional Application No. 60/072,455, January 26, 1998. This choice in the range of 300-420 nm reduces or eliminates the need for additional filters between the tissue and the imaging detector

due to the fact that standard, electronic imaging sensors, as used in video endoscopes, are insensitive to the excitation light. Visible red light, to which the imaging detector is very sensitive, is chosen to illuminate the tissue for the purpose of acquiring a reference image. This reference light is passed through the same optical guide as the excitation light, and illuminates the tissue with the same normalized spatial distribution and angular distribution as the excitation light. By correlating the spatial intensity and angular distribution of light used for the reflected image and the light for the fluorescence image, a more accurate and diagnostically useful imaging system is provided. This system of the present invention allows the reference image to be used to normalize the fluorescence image so that local reductions in fluorescence intensity can be accurately quantified. In another preferred embodiment, using color wheel (monochrome CCD) video endoscopes, the reference light and excitation light are applied sequentially. In another preferred embodiment, using color CCD video endoscopes, the fluorescence excitation light and reference light are applied simultaneously.

An additional reference image can be taken using a third color channel of the video colonoscope. The reference light directed onto the tissue for this image is predominately blue in color, with a central wavelength and bandwidth which approximates the fluorescence spectrum induced by the UV excitation light. The ratio B/R of this blue reference image, B , to the red reference image, R , on a pixel-by-pixel basis, identifies areas of the imaged tissue where chemical species (predominately hemoglobin) absorb some of the intrinsic fluorescence before it can exit the tissue and be imaged. The acquisition of the ratio F/R of the fluorescence image, F , to the red reference image, R , consequently provides two parameters, F/R and B/R , that can be used on a pixel-by-pixel basis for determining whether abnormal tissue such as dysplasia is present in the imaged region of interest.

BRIEF DESCRIPTION OF THE DRAWINGS

Figure 1 is an illustration of the fluorescence spectrum of normal and dysplastic colon tissue due to excitation with UV light at 370 nm.

Figures 2a and 2b are diagrams of the fiberoptic delivery probe system in
5 accordance with the present invention.

Figure 3 is a schematic diagram of an embodiment of the fluorescence imaging endoscope of the present invention which uses a separate fiberoptic light guide to deliver the UV excitation light through a biopsy channel of a color-wheel (monochrome CCD) video endoscope.

10 Figure 4 is a flow chart for the production of the output image.

Figure 5 illustrates a timing diagram for the image acquisition, analysis and display process in accordance with the present invention.

Figures 6a - 6c illustrates the UV output from a mercury arc lamp as a function of pulsed current in accordance with the present invention.

15 Figure 7 is a schematic diagram of a computer-controlled, pulsed lamp power supply in accordance with the present invention.

Figures 8a - 8d illustrate the details of opto-mechanical elements of the pulsed light source in accordance with the present invention.

Figures 9a - 9c illustrate the construction of the optical delivery fiber used in
20 accordance with the present invention.

Figures 10a - 10b are schematic diagrams of a preferred embodiment of the imaging system which uses a separate fiberoptic light guide to deliver the UV excitation light through a biopsy channel of a color CCD video endoscope.

Figure 11a - 11d are schematic diagrams of a preferred embodiment of the
25 imaging system which integrates the UV-transmissive illumination guide into the endoscope and connects an external UV excitation source to a standard white light source with an adapter module.

Figure 12 is a schematic diagram of a preferred embodiment of the imaging system which fully integrates the excitation light source and white light source into

an illumination system for an endoscope fitted with UV-transmissive illumination guide.

Figures 13a and 13b illustrate another preferred embodiment of an illumination and collection endoscope imaging system.

5 The foregoing and other objects, features and advantages of the invention will be apparent from the following more particular description of preferred embodiments of the invention, as illustrated in the accompanying drawings in which like reference characters refer to the same parts throughout the different views. The drawings are not necessarily to scale, emphasis instead being placed upon
10 illustrating the principles of the invention.

DETAILED DESCRIPTION OF THE INVENTION

The present invention relates to systems and methods for imaging autofluorescence from epithelial tissue to highlight regions of dysplasia. The tissue autofluorescence imaging system in accordance with the present invention improves
15 upon the previous state of the art in that, in its simplest form, it can be added on to an existing video endoscope system with no modifications to the endoscope itself, and with only the addition of a shutter in the optical path of the endoscope's visible light source. The handling characteristics of the endoscope are thus not adversely affected by the addition of image intensifiers and external cameras at the proximal
20 end as required by fluorescence imaging systems currently on the market. Switching between normal visible light imaging, in full color, and fluorescence imaging is accomplished by an electronic switch rather than a physical manipulation by the clinician, also required by current systems. The resulting fluorescence video image is processed by a computer so that the diagnostic image seen by the clinician
25 consists of a familiar visible light image (in gray scale) with a false color overlay indicating those areas in the image where the fluorescence from the tissue is reduced in comparison to the fluorescence from normal tissue. This image is much easier to interpret than the combined red/green raw fluorescence images provided by current systems, particularly for red/green color-blind clinicians.

The system of the present invention uses only a single imaging detector at the distal end of the endoscope for acquiring normal color images, fluorescence images and visible reference images. The use of a camera at the distal end is made possible by using fluorescence excitation light at ultraviolet to deep violet
5 wavelengths to which the CCD camera is insensitive or can be made insensitive using a fixed filter. This allows broadband collection of the resulting tissue autofluorescence from blue to red wavelengths, resulting in sufficient light for effective video signals without the need for additional image intensification. Fluorescence imaging in this fashion has been described *in-vivo* by Wang, et al. as
10 described in U.S. Application No. 09/238,664, entitled "Fluorescence Imaging Endoscope" filed on January 26th, 1999, the entire contents of which is incorporated herein by reference.

Previous autofluorescence imaging systems have depended upon imaging very weak fluorescence at red wavelengths to provide an image to compare with the
15 fluorescence image taken at blue-green wavelengths. Additional image intensification is particularly necessary to provide a usable red fluorescence images. The autofluorescence imaging system and method of the present invention avoids this expense and difficulty by supplying additional visible red light illumination to the tissue for the purpose of obtaining a reference image. To be effective, however,
20 the UV excitation light and the visible reference light must be delivered to the tissue through a common optical guide and exit the same illumination aperture with the same angular distribution. This requires a careful design of the excitation and reference light source optics. The processing of the reference image includes other features, such as histogram analysis, to eliminate artifacts such as visible specular
25 reflections which do not occur in fluorescence images.

The combination of the system features in accordance with the present invention can be realized in several different ways, depending upon the endoscope with which the autofluorescence imaging system is combined. Some of these features can also be built into the endoscope itself in stages as the technology is
30 accepted by clinicians.

In a preferred embodiment, the excitation illumination and reference illumination are generated by the same arc lamp source and delivered to the tissue through the same fiberoptic probe passed through a biopsy channel of the endoscope. In other embodiments the excitation light can be generated with a stand-alone source but delivered to the tissue through UV-transmissive illumination bundles built into the endoscopes as replacements to the standard, UV-absorbing glass bundles. The reference illumination is derived from the normal white light illumination source by switching in a red-pass filter that absorbs light at blue and green wavelengths which can interfere with imaging the predominately blue fluorescence. A deep red light source or near infrared (e.g., at 670 nm or above) can be used without affecting the normal visible illumination and can be combined with a dichroic element for delivery along a common optical path. In further embodiments, the excitation light source can be combined with the standard, white light illumination source and delivered to the tissue through an endoscope with UV-transmissive fiberoptic illumination bundles. A rotating wheel light source can also be modified to produce four colors, ultraviolet, blue, green and red, to perform both the fluorescence imaging and the visual imaging.

Figures 2a and 2b show the general components of a preferred embodiment in which the excitation light and the visible reference light are delivered to the tissue through a separate fiberoptic probe. The fiberoptic probe 200 is passed through a biopsy channel of a standard video endoscope 202 with its tip finally placed at or near the distal tip of the endoscope 204. When a footswitch 206 is depressed by the clinician, the normal white light illumination from the endoscope's light source and video processor 208 is switched off with a shutter. This white light normally illuminates the tissue through the two fiberoptic illumination ports 210 and 212 on the distal tip of the endoscope. Simultaneously, a complementary shutter in the excitation and reference illumination source 214 is switched on so that excitation and reference light can pass through the fiberoptic probe 200. The excitation and reference light exits the tip of the fiberoptic probe 216 and illuminates the tissue 218. The video image detection system 220 transmits the resulting fluorescence

image signal and reference image signal back through the endoscope 202 to the video processor 208 where they are converted into different color channels R and B) in a standard R,G,B National Television Standards Committee (NTSC) video format. These two channels are digitized with a video framegrabber in the computer system 222. The digitized fluorescence and reference images are processed together in real time to quantify image regions where the fluorescence is reduced compared to normal tissue. Reduced fluorescence is the primary indicator of dysplasia. Areas of the tissue which are likely to be dysplastic are highlighted with false color in a processed image of the tissue which is displayed on a computer monitor 224 and updated at a rate of up to 10 Hz. The preferred embodiment shown in Figures 3a and 3b is thus an add-on component to an existing endoscope/ video processor system which only requires the addition of an internal shutter to the white light source in the endoscope systems video processor. For color-wheel (monochrome CCD) video endoscopes, the excitation light source 214 is constructed to provide sequential excitation and reference illumination as described in more detail hereinafter. For color-CCD video endoscopes the excitation light source 214 provides simultaneous excitation and reference illumination, also described in more detail hereinafter.

The autofluorescence imaging system demonstrated by Wang, et al. used an argon-ion laser as a UV excitation source. Other laser sources can be used including solid state lasers, such as gallium nitride laser diodes, operating at wavelengths in the range of 380 nm to 420 nm, which have smaller size and low power operation. The systems in accordance with the present invention uses a mercury arc lamp as a source of UV excitation with a spectral band around the 365 nm mercury line. The mercury arc source is smaller, and less expensive than the argon-ion laser, requires relatively little power and is air-cooled. For the autofluorescence imaging systems used with color wheel (monochrome CCD) video endoscopes, the current to the arc lamp can be pulsed. In the color wheel video systems, the normal light source provides red, green and blue light pulses sequentially during a 33 ms video frame which are combined with the video processor to provide a color image. In the autofluorescence imaging mode these systems replace the normal blue light pulse

with a UV pulse and the green light pulse with a (nominally red) reference light pulse transmitted through the same optical fiber as the UV pulse. By pulsing the arc lamp current for the 8 ms period that the blue light is normally on, the UV source can provide as much or more excitation fluence to the tissue than it could if the CW lamp intensity was integrated over the full 33 ms video frame period.

With both color wheel video systems and color CCD video systems the autofluorescence imaging mode is initiated with a footswitch which controls complementary shutters on the excitation/ reference light source and the normal endoscope white light source. Shutting off the normal endoscope illumination is required in order to image the weak fluorescence from the tissue. Two different autofluorescence imaging modes can be used. In one mode, a single frame of the false color overlay from the processed autofluorescence image is combined with the immediately preceding (or following) color image and frozen on the computer screen. In the second mode the processed fluorescence image is updated continuously (at about 7.5 Hz to 10 Hz allowing for the image processing time) for as long as the footswitch is depressed. In the continuous operation mode the visible image is shown as a gray scale (since it is taken with a monochrome reference illumination) with false color overlays showing regions of likely dysplasia.

Figure 3 shows an overall schematic diagram of a preferred embodiment of the fluorescence imaging system for use with color wheel (monochrome CCD) video endoscopes. For normal mode of operation of the color wheel video endoscope, the distal tip of the endoscope 300 is inserted into the hollow organ of the body to observe an area of tissue 302 which may contain a region which is dysplastic 304. The CCD imaging device and lens subsystem (the video camera) 306 is flanked on either side by the normal illumination apertures 308 and 310.

The normal illumination light is carried by a fiberoptic bundle 312 which extends through the length of the endoscope and bifurcates near its distal tip, terminating at apertures 308 and 310. A shutter 314 between the illumination source and the fiberoptic bundle, controlled by a digital signal 315, allows the white light illumination to be turned off without turning off the source lamp 316. In the type of

endoscope shown, the normal color image is obtained by combining three, successive images taken with red, green and blue light pulses provided by a rotating filter wheel 318. The CCD detector in this type of endoscope is sensitive to all wavelengths between 400 nm and 700 nm, but is insensitive to UV excitation wavelengths around 365 nm which are used to excite the autofluorescence. This is due both to the design of the silicon sensor array and to the choice of optical materials used to physically protect the surface of the array. The CCD detector continuously integrates all of the light which falls on its surface so the illumination must be shuttered off while the CCD rows are shifted down to the readout electronics or else a streaking effect will be seen on the image. The red, green and blue light pulses have a duration of about 6 ms followed by a dark period of 5 ms during which the camera pixels are read out, yielding a total video frame period of about 33 ms or 29.97 frames per second to meet the NTSC standard. The analog readout signal from the CCD camera is carried through a cable 320 to the endoscope video processor 322.

The three successive monochrome images are digitized and combined into a standard color video signal at the end of the video frame. The processor has two groups of standard red, green, blue (RGB) plus synchronization outputs. One output group of color signals 324 goes to the endoscopes video monitor 326 to display the normal color image of the tissue 328. Another group of color signals 330 goes to a video framegrabber 332 in the fluorescence imaging computer system 334 which will acquire and process the autofluorescence and reference images. A standard composite color signal output 336 from the video processor goes to a synchronization circuit 338, based on a National LM1881 Video Sync Separator. This synchronization circuit 338 determines when the interlaced even and odd fields occur in the video signal and outputs a binary digital signal 340 which is high during the odd field and low during the even field. This signal 340 is used throughout the fluorescence imaging system to synchronize its functions to the timing set by the endoscope's video processor.

To engage the autofluorescence imaging mode of operation, the clinician depresses footswitch 341 which sends a signal on cable 342 to the computer 334. At the appropriate time, synchronized to the next opportunity as determined by the synchronization signal 340, a signal is sent out on the shutter trigger line 315 to shutter 314 on the normal illumination source and the shutter 343 on the excitation/
5 reference light source. These shutters are complementary, so that the signal on line 315 opens shutter 343 and simultaneously shuts shutter 314. The excitation and reference light pulses, generated by the rotating shutter wheels in the optical group 350, then pass into the fiberoptic probe 344. Probe 344 is inserted into a biopsy
10 channel opening 345 and slid down the channel until its end window 346 is either at or just beyond the distal tip of the endoscope 300. The excitation and reference light pulses illuminate the central portion of the endoscope's visual field. The angular extent of the illumination depends on optical elements 350, the fiber numerical aperture, and the optical characteristics of the fiberoptic probe end window 346.

15 The excitation and reference light pulses must occur during two of the three normal illumination pulse periods for the autofluorescence and reference images to be properly acquired. These images appear on the next video output frame on two of the three video output channels from the endoscope video processor. The appropriate timing is accomplished by rotating shutters 352 and 354 in the excitation (UV) and
20 reference (RED) light paths, respectively, of the optical group 350. The shutters are driven by direct current (DC) motors 356 and 358 which rotate at a speed which can be controlled by varying their supply voltage. A fiducial hole near the rim of the shutters, combined with an optical source and detector, generates a reference pulse for each which marks its phase as it rotates. A phase-locked loop (PLL) 362 and 360
25 for each motor, 358 and 356, respectively, adjusts the motors voltage so that the reference pulse for each matches the rising edge of the synchronization pulse 340 marking the beginning of the odd video field.

By positioning the hole in the shutters appropriately the excitation light pulse and the reference light pulse can be timed to the exposure periods of the camera. The
30 excitation pulse is timed to match the normal blue exposure since this exposure

period is somewhat longer (8.1 ms) than the others. The reference light pulse is timed to match the normal green light exposure (5 ms) since it is the next longest of the normal exposure periods. The normal red exposure period, which immediately follows the normal blue exposure period, is not currently used, but can be employed to obtain additional autofluorescence or reference images or an additional visible reflectance image for other spectroscopic analyses. During the excitation exposure the current to the mercury lamp can be boosted to a higher level to increase the excitation light output. The lamp power supply utilizes a DC current section 366 to maintain the idle current and start the lamp.

10 A computer-controlled, pulsed current section 368, connected in parallel with the DC section 366, can rapidly switch in multiple, parallel-coupled, constant-current sources to vary the output power of the lamp as required by the imaging system. The current pulses are synchronized to the video system using the same synchronization pulse 340 that serves to lock the rotating shutters. The computer digital input/output (I/O) section 370 outputs a digital pulse 372 which is combined with the timing pulse 340 to boost the lamp current during the excitation exposure. The number of constant current sections which are triggered in parallel can be varied by a set of digital control lines 374 as required. If the tissue is sufficiently close to the excitation/ reference delivery probe window 346 then no boost may be required. If the peak value of the autofluorescence image drops to a minimum acceptable level, as determined by the computer image analysis program 376 as monitored by the overall control program 378, then additional current boost sections are enabled as required for the next exposure.

25 Once the autofluorescence image and the reference image have been acquired by the digital framegrabber in the computer then the analysis can begin. A reflectance (non-fluorescing) image taken with an endoscope camera system measures the brightness of the tissue surface in its field of view. To the extent that the tissue surface is a Lambertian (non-specular) reflector, the reflectance image signal (as digitized by the video framegrabber for each discrete pixel in the video image) is proportional to distance of the tissue from a single illumination source (or

a weighted distance from multiple sources), and to the integrated energy of the excitation illumination during that video frame. The end window 346 of the excitation/ reference delivery probe is not in the direct line-of-sight from the camera lens to the tissue so there will be visible shadows. A reflectance image can thus be used to measure the excitation illumination at the tissue surface visible to the camera, including the presence of shadows in the autofluorescence image. Note that this is only true if the excitation illumination and the visible illumination emanate from the same aperture with the same transverse intensity profile and angular divergence as provided by the design of the source as described herein above. A visible reflectance image taken with light from the illumination bundles of a standard endoscope, for instance, is not acceptable for determining the excitation illumination from a separate optical fiber passed through a biopsy channel of that endoscope. Note also, however, that the same aperture/ same divergence condition can be satisfied by passing both the excitation and the reference light pulses through the illumination bundle of the endoscope.

The sequence of steps for obtaining a false color indication of dysplasia probability using the visible reference image along with the autofluorescence image proceeds as follows. The two images are first corrected for the gamma factor applied to the video signal by the video processor which is generally connected to a video monitor rather than a framegrabber. This insures that the digitized images acquired by the framegrabber in the computer are linear functions of the illumination fluence (time-integrated intensity). The two images are then normalized to their peaks, which is generally a region of non-dysplastic tissue somewhere in the visual field. There are a few pixels in the reference image which are saturated due to specular reflections of the reference illumination. These are effectively eliminated by producing a histogram of the reference image and normalizing the image to a peak value. This generally includes about 99% of the pixels.

All normalized reference image pixels above a value of 1 (the specular reflections) are then reset to a value of 1. There are no specular reflections in the autofluorescence image so such a histogram-based normalization is not necessary.

On a pixel-by-pixel basis the autofluorescence image value is then divided by the corrected reference image value to produce a ratio image. This division is performed only when the autofluorescence and reference image pixel values are above a minimum threshold condition to insure that the analysis is not attempted with too little illumination to provide a reliable measurement.

If a ratio image pixel value falls below a predetermined value (typically one-half to one-third) then that pixel represents a region of reduced fluorescence on the tissue surface which is indicative of dysplasia. The corresponding pixel in the processed output image can then be set to a false color state to indicate the relative probability of dysplasia. If the ratio image pixel value is below $1/3$ the red value of the corresponding processed output image pixel it is set to the value of the reference image and the green and blue values of that pixel are set to zero (it is a shaded red, indicating a high probability of dysplasia). If the ratio image pixel value is $1/2$ to $1/3$, then the green value of the processed output image pixel it is set to the reference image value and the red and blue values are set to zero (it is a shaded green, indicating a moderate probability of dysplasia). If the ratio image pixel value is above $1/2$ then the red, green and blue values of the processed output image pixel are all set to the reference image value (it is a shaded gray, indicating the probability that the tissue is normal). This processed output image 386 is displayed on an LCD monitor 384 attached to the system computer 334.

Figure 4 illustrates a process sequence 400 after a patient has been prepared the endoscope inserted into the body cavity or human and the distal end thereof has been positioned for imaging of a region of interest. In this particular example, a visible reference image is obtained 402. This reference data is corrected 404, a histogram is generated 406, the data is normalized 408, selected pixels are reset 410 and a threshold value is applied 412. After a fluorescence image is obtained 420, the image is corrected 422, normalized 424, a threshold applied 426, and a ratio image is generated 430. The resulting output image or representation is then compared 432 to a reference, and a given region is determined to be normal 440 or dysplastic 450.

Figure 5 shows a timing diagram for one cycle of the imaging process. The maximum video output rate is 29.97 Hz as set by the NTSC standard. The diagram indicates a system which requires one frame time for acquisition, between two and three frame times for analysis, and a fraction of a frame time for transferring the result to the output image buffer. The resulting update rate for the analysis image is 7.5 frames per second. With a single, fast processor, the analysis time can be reduced and the update rate correspondingly increased to 10 or 15 frames per second. The image analysis can also be performed at output rates up to 30 frames per second by, for example, using two processors operating in parallel with only a few frames of delay between acquisition and display. Increased update rates, however, require a reduced upper limit on the pulsed current to the lamp in order to maintain an average lamp dissipation of no more than 100 W. Figure 5 also shows how the mercury lamp current is boosted only during the blue exposure periods which result in autofluorescence images. At other times the lamp idles at a reduced power.

Figures 6a and 6b show that the UV output power from a 100 W Hg arc lamp, pulsed for 8 ms above a 70% idle current, is essentially a linear function of its input power to at least a factor of 5 over its nominal rated power. Boosting the lamp current during the UV exposure period increases the lamp output power which, in turn, allows a larger area of tissue to be scanned for dysplasia. It also provides a means for adjusting the output of the lamp for an optimum video exposure on a pulse-to-pulse basis. Since the lamp discharge maintains a nearly constant voltage drop across the arc regardless of current, the lamp output power is essentially proportional to current. About 70% power to the lamp must always be maintained, however, to keep the mercury in the vapor phase. Figure 6c shows that pulsing the current to a factor of 5 over the rated CW current can be repeated continuously at a rate of 7.5 Hz.

Figure 7 shows a block diagram of the lamp power supply as it connects to the mercury arc lamp 700. The DC idle current circuit 702 and high voltage (HV) starting pulse circuit 704 are handled by a single commercial power supply designed

for CW 100 W mercury arc lamps. The constant current, triggered pulse circuit 706 is designed to match the specific requirements of the fluorescence imaging endoscope. Four of these circuits are connected in parallel so that the output of the arc lamp can be set digitally to one of 5 different power levels (including idle). Each circuit is composed of a power MOSFET switch which adjusts its resistance to keep the passed current at a fixed level, typically equivalent to the normal DC current of 4 amps. These circuits can be individually triggered by a computer whenever their current boost is required to properly illuminate the tissue. Each circuit draws its current from a storage capacitor 708 charged by a power supply 710 through a current limiter 712. This limited current design minimizes the possibility of fault conditions which could overload the power handling capacity of the arc lamp.

Figures 8a, 8b, 8c and 8d show details of the excitation/ reference illumination source optics. A single mercury arc lamp 800 in Figure 8a is used as the source of both wavelengths because both the UV/violet excitation light and the red/near infrared reference light need to have the same apparent source volume 802 at the distal tip of the endoscope. Small 100 W, mercury arc lamps have arc dimensions of 0.5 to 1.0 mm which are small enough to couple efficiently into the optical fibers used to deliver the light to the tip of the endoscope. Separate lamps can be used for the excitation and reference beams, but the tolerances for the source optics and the alignment of the source arcs are then more critical. In the preferred, single lamp, source design shown in Figure 8a the light collection optic 804 is shown schematically as a single, fused silica, UV-transmitting lens. In practice either a multiple element lens design or a mirror-based Schwarzschild objective is used to reduce optical aberrations on the collected beam. Such collection optics could also collimate light from the focal volume of a commercial mercury arc lamp with a fixed, pre-aligned, elliptical reflector as an alternative source. The collection optics 804 collimate the light from the lamp so that it can be effectively filtered into UV and visible components by means of a dichroic mirror 806 which reflects UV and/or deep violet wavelengths and transmits visible wavelengths.

Separation into two paths is required so that two rotating shutters, 808 and 810, can produce UV excitation and visible reference pulses at different times during the video frame as described above. A UV-reflective surface coating on mirror 812 immediately after the dichroic mirror reflects nearly 100% of the desired UV light in that path and dumps much of the undesired visible light reflected from the dichroic into its substrate. Additional UV filters represented by element 814 may include absorptive filters such as Schott UG-1 glass as well as multilayer dielectric bandpass filters centered on the 365 nm line of mercury. The UV path rejects visible light to a high degree since the efficiency of the 460 nm tissue fluorescence is only about 0.1%. Leakage of visible source light during the UV exposure period reduces the contrast of the autofluorescence image. Some correction for such leakage is possible during the digital image processing but corrections always add a small amount of noise to the result.

The first element in the reference path is a weak lens 816 to correct for the lower refractive index of the fused silica collimating lens 804 at visible wavelengths as compared to the UV. Note that lens 816 would not be necessary if a Schwarzschild objective is used at the position of lens 804 since such a design uses only mirrors and is thus totally achromatic. An attenuator can optionally be moved into the reference path at position 817 to prevent saturation of the reference image at close inspection distances. This attenuator must be uniform across the area that the reference beam passes through in order to maintain the uniform angular intensity distribution required of the reference illumination. This attenuation can be accomplished with variable, crossed polarizers, differentially-sliding, linear attenuation wedges or electromechanically-switched, fixed-value attenuators. The visible filter 818 in the reference beam path is less critical than the UV filter when sequential excitation/ reference beams are used such as in the system shown. The reference beam wavelength is chosen to avoid hemoglobin absorption bands since a significant absorption introduces an error into the analysis which assumes that the reference image is equivalent to a measurement of the excitation illumination intensity. The rotating shutters 808 and 810 follow the filters.

Continuing down the visible path after shutter 808 is turning mirror 820 with a broadband visible reflective coating. A second dichroic mirror 822, identical to dichroic mirror 806, recombines the UV excitation beam and the visible reference beam onto a common path. One additional turning mirror 824 which reflects both

5 UV and visible light directs the two beams towards the focusing optics 826 which couples them into the delivery fiber 828. The turning mirror 824 makes the number of reflections in both the UV and visible paths of the system equal. With equal numbers of reflections, any change in the position of the mercury arc 802 relative to the position of the collection optics 804 results in the same angular deviation for

10 both the excitation and reference beams. Equal deviation angles preserves the overlap of the excitation and reference beams on the tissue. The system as shown in Figures 8a - 8c, in which the combined beams exit the optical train in the same direction as the incoming beam, also makes the direction of the output beams invariant under translations and small rotations of the optical train as a whole.

15 The aperture stop 830 in Figure 8a insures that the UV excitation and visible reference beams have the same angular convergence on entering the delivery fiber 828. The transverse dimensions of the two beams at the position of the focusing optic 826 will invariably be slightly different due to small errors in the positioning of the optics in the two paths and slight differences in the effective emitting volume

20 of the arc at the two wavelengths. The aperture stop 830 is set so that both beams are slightly clipped at their outside edge, insuring that the maximum angular input to the optical fiber is the same for both. Figure 8b shows that, to a great degree, the input angle to an optical fiber, relative to the fiber axis, is preserved upon each reflection within the fiber. Illumination of a fiber with a collimated beam from a single

25 direction generally results in a cone of emitted light from the opposite end of the fiber which has the same angle relative to the axis. The light is spatially averaged over the exit aperture of the fiber but the angles only slowly spread due to bends in the fiber as it passes down the length of the endoscope. The embodiment shown insures that the normalized angular intensity distribution of the excitation

30 illumination and the reference illumination are closely matched as required by the

autofluorescence normalization method. If, for example, the normalized reference illumination locally exceeds the normalized excitation illumination in a region 831 as shown in Figure 8c, then the analysis falsely indicates reduced autofluorescence in this region. If the normalized reference illumination is greater by a factor of two over the normalized excitation then the analysis provides a false positive indication of dysplasia, where the threshold condition is set to 50%. Similarly, if the normalized reference illumination is a factor of two below the normalized excitation illumination in some region, then the analysis indicates normal tissue even if the autofluorescence was actually reduced by the same factor of two in that region.

Such large errors are unlikely in the center of the illumination field but can occur at the edges of the field where the illumination drops to the noise level. The threshold conditions placed on both the reference and autofluorescence images insure that the analysis is not attempted where it is too easily influenced by noise or edge effects. In general, it is described to maintain any variation in the intensity distribution of the reference light as seen at 860 (normalized) at less than 20% relative to the intensity distribution of the excitation light at any point along the optical path between the combiner and the tissue surface. It is most important to achieve minimal variation in intensity at the tissue surface. Thus light exiting the fiber 862 is within the cone defined by the angle θ and variations in intensity at along the wavefront 864 should be less than 20% to minimize the risk of false image.

Figure 8d shows a more detailed view of the rotating shutter design in the excitation/ reference source. Each blade rotates once in 33.3 ms, the period of one video frame. The angle subtended by the aperture 832 on the excitation shutter blade 810 corresponds to the 8.1 ms period for the normal blue exposure. The timing of the UV exposure relative to the video frame is set by the phase-locked loop matching the rising edge of the optical pulse through the excitation blade fiducial 834 to the rising edge of the odd field mark on the synchronization signal 340 as shown in Figure 3. The apparent position of the excitation beam in the optical train is marked by the dotted circle 836. For the relative dimensions of the beam diameter to shutter diameter shown the total rise time of the light pulse is 1.7 ms which is

acceptable compared to the total length of the pulse. The blade 810 is thin, light in weight and opaque, made from a material such as stainless steel or glass-fiber reinforced epoxy.

Minimizing the moment of inertia simplifies locking the shutter's rotation rate with the PLL. Being opaque is important to prevent light leakage. The cutouts 838 around the central hub 840 of the wheels leave thin blade flexures 842 which allow the wheel to flex slightly as it rotates to stay perpendicular to the rotation axis of the motor, minimizing vibration and stress on the motor bearings. The continuous outer edge of the shutter serves to carry the timing fiducial at the maximum distance from the axis where it is most sensitive and serves a safety function as well in the case of an inadvertent touch during assembly and testing. The aperture in the visible reference shutter 844 is positioned to match the normal green or red period of the video timing sequence. This instrument can either provide a single false color image for each footswitch signal or a series of processed images at a submultiple of the 30 Hz video frame rate.

A standard personal computer system can process the acquired data and update the false color image at 10 frames per second which is fast enough to preserve the impression of smooth motion and slow enough to allow a significant increase in the pulsed lamp current. Operation at lower submultiples of 30 Hz (7.5 or 6 Hz) still provides adequate real-time feedback and permits higher pulsed currents (and thus brighter UV illumination) without exceeding the nominal 100W average power dissipation in the lamp.

Figures 9a - 9c show details of the optical fiber probe which delivers the UV and reference light pulses from the source described above, through a biopsy channel in the endoscope, to the tissue at the distal tip of the endoscope. The delivery fiber 900 must transmit both the UV and the reference light wavelengths, couple efficiently to the light source and be sufficiently flexible so that it doesn't significantly affect the flexibility of the endoscope tip.

The fibers can be made with multiple, small-diameter fused silica fibers, but the preferred system uses a single, UV-transmitting acrylic fiber with a diameter of

from 1.5 to 2 mm. One particular acrylic fiber that is suitable is the Raytela Polymer Optical Fiber made by Toray Industries, Inc. The single fiber system increases coupling efficiency by eliminating the packing fraction losses of multifiber bundles. Eliminating the multiplicity of internal cavities also increases the reliability of the disinfection process between procedures.

A window 902 at the tip of the probe is held on with a sleeve 904 and low-fluorescence epoxy in the gaps 906, 908 and 910 as shown in Figure 9b. A thin, bio-compatible, sheath of heat-shrink tubing protects the thin cladding of the optical fiber. A band of epoxy at position 914 in the gaps between the fiber 900, the sheath 912 and the sleeve 904 seals the probe for disinfection and holds the sheath in place. This type of fiber has a numerical aperture, NA, of 0.5 which means that the light is emitted over a 60 degree, full-angle cone. Probes with a plane window illuminate only about half of the 120 degree maximum field of view for a typical endoscope.

Figure 9c shows a negative lens 916 used in place of the plane window 902 with an air gap 918 between the lens and the end of the optical fiber. The negative lens increases the illuminated field of view at the expense of excitation illumination intensity. Additional negative lenses would increase the field further. Probe designs such as that of Figure 9c are optimized for scanning large areas while the plane window design of Figure 9b is optimized for scanning more localized regions of dysplasia. These windows or lenses can be made of fused silica or UBK-7 to optimize UV transmission. Millimeter thicknesses of common glasses such as BK7, however, do not absorb a significant fraction of the excitation illumination. The lenses or windows can also be made from a UV-transmitting, blue-blocking, red-transmitting glass such as Schott UG-1 if the level of blue fluorescence in a particular type of delivery fiber degrades the fluorescence image. Such additional filtration has not been required in probe designs using the Raytela fiber.

In a preferred embodiment, constructing the above probe from a UV-transmissive plastic will reduce the cost, compared to a fused silica construction, to the point where the entire probe can be disposed after a single use.

Figures 10a and 10b show how the autofluorescence imaging system shown in Figure 3 can be modified for use with a standard, color-CCD video endoscope 1000. In this type of endoscope the image detection system 1002 detects red, green and blue light simultaneously with discretely filtered pixels on the CCD detector.

- 5 The illumination source 1008 for this type of video system emits continuous, broadband, white light which still needs to be turned on and off with a shutter 1010 for use with the modified autofluorescence imaging system. In this embodiment the excitation/ reference light source 1018 produces the two wavelengths simultaneously with the optical train 1020, with a high rejection of light at blue and green
- 10 wavelengths (greater than 1000:1) where the autofluorescence peaks. The spectrum of the continuous illumination is shown in Figure 10b.

- For autofluorescence detection the complementary shutters 1010 and 1024 are triggered as before, causing the tissue to be illuminated with both the excitation light and reference light simultaneously. The UV-induced autofluorescence
- 15 (primarily at 460 nm) is then detected by the blue-responsive pixels in the CCD camera. The red reference reflectance image is simultaneously detected by the red-responsive pixels. Color CCD cameras typically use electronic shuttering so that they don't need a dark period for their readout. In this embodiment, the rotating shutters in the excitation/ reference light source 1018 are either stopped in an open
- 20 position with a detent, or eliminated altogether if the autofluorescence system is to be used only with color video endoscopes. The UV illumination thus lasts for the full 33 ms of the frame, increasing the integrated visible fluorescence and reference signals and eliminating the need for a pulsed current supply. The lamp current supply 1022 could still be pulsed occasionally for single images without overheating
- 25 the lamp. The analysis would proceed as before with the reference image appearing on the red channel of the RGB NTSC signal and the autofluorescence image appearing on the blue channel.

- In the fluorescence image endoscope in accordance with the present invention, both the UV-excitation light pulse and the visible-reference light pulse
- 30 are delivered to the tissue through an optical probe inserted through a biopsy channel

of a standard endoscope. The excitation and reference light could alternatively be passed through the illumination bundle of an endoscope if that bundle were to be modified to transmit UV as well as visible wavelengths. The requirement would remain that the excitation light and the reference light, even if generated by different
5 sources, have the same angular distribution as they exit the distal tip of the endoscope.

Figures 11a - 11d show an embodiment of the autofluorescence imaging system which uses a UV-transmissive endoscope, a standard white light source for normal imaging, a separate excitation light source and a coupling box. This system
10 is used if a modified endoscope were to be developed before a dedicated light source combining the white light, excitation and reference functions in one unit. In this intermediate instrument, the adapter box 1100 would fit between the standard white light illuminator 1102 and the electrical/ optical connection plug 1104 of the UV-transmissive video endoscope. The optical connection plug 1106 normally fits
15 directly into the visible light source 1102 to collect the illumination light and pass it into the illumination bundle of the endoscope. In this embodiment the optical connection plug 1106 would fit instead into the adapter 1100 and an identical optical connection plug 1108 on the opposite end of the adapter would fit into the visible light source 1102. A set of imaging optics 1110 would transfer the light emitted
20 from the exit aperture of plug 1108 to the entrance aperture of plug 1106. This set of imaging optics would also transfer excitation light from a separate light source, exiting the optical fiber 1112 to the entrance aperture of plug 1106.

The imaging optics set is detailed in Figure 11b. The combination of the excitation illumination onto a common axis with the visible illumination is
25 accomplished with a dichroic beamsplitter cube 1114. An achromat 1118 collimates visible light expanding from a point 1118 on the exit aperture of plug 1108. A second achromat 1120 refocuses this light onto point 1122 on the entrance aperture of plug 1106. A set of fused silica lenses 1124 and 1126 nominally collimate the light from point 1128 on the exit aperture of the excitation delivery fiber 1112. The
30 position of these lenses is actually adjusted to provide the best imaging of point

1128 onto point 1122 since the achromat 1120 will not be corrected for UV wavelengths.

In this embodiment the reference illumination is derived from the normal red illumination light from the visible light source. Filter 1130 is made of red absorbing
5 glass which strongly attenuates wavelengths UV, blue and green wavelengths. Filter 1130 is mounted on a sliding base 1134, shown in Figure 11c, and is switched into the optical path electromechanically when autofluorescence imaging is to be used. The shutter in the UV light source is opened at the same time. The tissue will then be illuminated by both the UV and the reference light either simultaneously if a
10 color video endoscope is being used or sequentially if a color wheel video endoscope is being used. Filter 1132 is clear glass which transmits all of the normal visible wavelengths. Filter 1132 is switched in electromechanically when normal illumination is to be used so that the focal point of lens 1116 remains constant.

For the endoscope illumination bundle 1136 to be both UV-transmissive,
15 flexible and long-lasting it must be made of fused silica optical fibers rather than regular glass fibers. The primary optical difference between the two materials, other than transparency, is that fused silica fibers generally have a lower numerical aperture than glass fibers. This means that light is collected and emitted at smaller angles from the axis in fused silica fibers. The illumination bundle bifurcates at
20 point 1138 and exits the optical ports 1140 and 1142 at the distal tip of the endoscope. The lens elements in the ports 1140 and 1142 must also be modified (made more negative) to maintain the same illumination angle for existing, glass fiber based endoscopes.

The video imaging detection system 1144 at the distal tip of the endoscope
25 transmits its signals back down the electrical wires 1146 within the endoscope to the electrical connector 1147 on the connector plug 1104. A mating connector on the adapter box 1100 collects these signals which are transferred across the adapter 1100 to a connector plug 1148 identical to plug 1147 which completes the electrical connections into the video processor. Analysis of the autofluorescence and reference
30 images then proceeds as before.

Once the UV-capable endoscope is available, the excitation source can be built into the endoscope light source and video processor so that the entire autofluorescence imaging capability is included in the endoscope system itself. This preferred embodiment is shown schematically in Figure 12. The electrical/ optical connection plug 1200 is mated directly to the excitation/ reference/ white light source and video processor 1202. This system operates exactly as the embodiment shown in Figures 11a - 11d, using a dichroic cube beam combiner system 1204 and a reference light filter system 1206, but without the necessity of an external adapter.

The overall reduction in visible blue-green autofluorescence is the primary indicator of tissue dysplasia for the imaging systems described to this point. These imaging systems characterize the autofluorescence reduction with a single parameter for every pixel in the tissue image. The specific causes of this reduction have been described by Zonios et al. in "Morphological Model of Human Colon Tissue Fluorescence," IEEE Trans. Biomed. Eng., 43(2), 113-122, 1996, incorporated herein by reference. The primary fluorophore is collagen, a protein in connective tissue, which is found within the thin mucosal layer and which is the dominant component of the submucosal layer. The autofluorescence in dysplastic tissue is reduced when cells in the mucosal layer enlarge and crowd out the collagen and when hemoglobin is added in the tissue due to increased vascularization. The hemoglobin absorbs some of the autofluorescence. The ability to characterize both the intrinsic fluorescence of the tissue (in the absence of absorption) and the level of hemoglobin in the tissue improves the ability of the imaging system to correctly determine the likelihood of dysplasia.

The imaging systems described above utilizes two of the three available imaging periods in the 33 ms acquisition period of a standard, color-wheel (monochrome CCD) video colonoscope. During one period, the ultraviolet excitation light illuminates the tissue and an autofluorescence image, F, is acquired. During a second, period visible red light illuminates the tissue and a reference image, R, is acquired which determines the intensity distribution of the ultraviolet illumination on the tissue. The ratio of these two images produces a single

parameter, F/R , for every pixel in the image which represents the relative level of overall autofluorescence at that point. The algorithm for highlighting portions of the image where dysplasia is likely is primarily one of assigning appropriate threshold values to the F/R parameter for each pixel.

5 The improved imaging system described below and shown in Figures 13a and 13b utilizes the third imaging period to acquire an additional visible reflectance image to allow for a more complete spectroscopic analysis. In this system predominately blue light is directed onto the tissue with a central wavelength and bandwidth which approximates the fluorescence spectrum induced by the UV
10 excitation light. As before, this blue light exits the same illumination ports and has the same angular intensity distribution as the ultraviolet excitation illumination and the visible red reference illumination. The resulting blue reflectance image, B , indicates regions where chemical species in the tissue absorb light at these blue wavelengths and thus areas of tissue where the intrinsic autofluorescence is reduced
15 by absorption processes. As before, the red reference image, R , is used as the measure of the illumination intensity of all three wavelengths at the tissue surface. The specific parameter to indicate (primarily hemoglobin) absorption on a pixel-by-pixel basis is thus the ratio B/R . The two parameters, F/R and B/R , are available on a pixel-by-pixel basis for determining that the imaged area of tissue indicates
20 dysplasia. Two basic methods can be used. In one, the F/R ratio can be corrected by the B/R parameter to more closely indicate the true value of the intrinsic fluorescence which produces a single, modified F/R parameter to which a threshold is applied as before. A second procedure uses a two-dimensional surface, defined by the F/R and B/R parameters, which can indicate the probability of dysplasia. In both
25 cases the threshold values for determining which are to be highlighted is determined by comparing measured values of F/R and B/R during clinical tests with the results of pathology reports on biopsy samples taken from those sites.

 The system in Figures 13a and 13b uses a mercury arc lamp source 1300 as before. The same optical elements 1302 and 1304 are used to collect and collimate
30 the light produced which includes both ultraviolet wavelengths and the full range of

visible wavelengths. The optical path of the ultraviolet illumination remains the same. A dichroic mirror 1306 reflects most of the UV light and some visible light to a UV-reflective turning mirror 1308. Again, most of the UV light is reflected while most of the visible light is absorbed into the substrate of 1308. A UV bandpass filter 5 1310 removes almost all of the remaining visible light so that only UV light reaches the timing wheel 1312. The timing wheel 1312 insures that the UV light which continues on is only passed during one exposure period for the video colonoscope. Another dichroic mirror 1314 and a total reflecting mirror (both UV and visible) 1318 direct the UV beam to the focusing lens 1318. An iris 1320 defines the angular 10 extent of the UV light entering the delivery fiber 1322. The delivery fiber 1322 carries all of the illumination wavelengths to the end of the colonoscope to illuminate the tissue.

The difference between the improved autofluorescence imaging system and the systems described previously lies in the path that the visible wavelengths take 15 through the system. Both blue and red wavelengths approach the visible timing wheel 1324. This wheel has two apertures which allow light to pass during the two remaining exposure periods for the video colonoscope. One of the wheel apertures is covered by a red filter 1326 which passes wavelengths appropriate for the reference image, R. These wavelengths are chosen to avoid both hemoglobin absorption bands 20 in the tissue and absorption bands in the optical fiber used to deliver the light to the tissue. The light to be filtered is well collimated so that multilayer dichroic coatings can be used, but simple absorption filters made of colored glass or plastic are adequate. The second aperture on timing wheel 1324 is covered by a broadband blue filter which approximates the spectrum of the intrinsic fluorescence produced by the 25 excitation light. The specific bandpass spectrum of this filter can contain wavelengths absorbed by hemoglobin and be sufficiently reproducible so that the resulting analysis program can be used by other instruments. The second aperture matches the third exposure period of the video colonoscope. Photoelectric detectors responding to light passing through timing holes 1330 and 1332 on the two wheels 30 as shown in Figure 13b provide electrical pulses for the phase-locked loops which

keep the timing wheels synchronized with the video acquisition system of the video colonoscope.

The remaining differences between the improved, three-color diagnostic system and the two-color system include a broadband, reflective mirror 1334 that must
5 reflects both blue and red visible wavelengths and an achromatic, long-focus lens 1336 which corrects for the longer focal length of the fused silica collection lens 1306 at visible wavelengths.

While this invention has been particularly shown and described with references to preferred embodiments thereof, it will be understood by those skilled
10 in the art that various changes in form and details may be made therein without departing from the scope of the invention encompassed by the appended claims.

CLAIMS

What is claimed is:

1. A fluorescence imaging system comprising:
 - a light source for producing excitation light that induces visible
 - 5 fluorescence in tissue and a reference light;
 - an optical combiner that combines said excitation light and said reference light onto a common optical path, said combined light being coupled into an optical guide that delivers the combined light to the tissue;
 - an image sensor that detects a fluorescence image and a reference
 - 10 image of the tissue; and
 - a data processor that processes the fluorescence image and said reference image to produce a processed output image of the tissue.
2. The system of Claim 1 wherein the light source is an arc lamp.
3. The system of Claim 2 wherein the arc lamp current source is a pulsed
- 15 source.
4. The system of Claim 1 wherein the optical guide is a removable fiberoptic extending through a biopsy channel of an endoscope.
5. The method of Claim 1 wherein the image sensor is located at a distal end of a endoscope.
- 20 6. The system of Claim 1 wherein the excitation light and the reference light are emitted sequentially such that a monochromatic image sensors detects a fluorescent image during a first time period and detects a reflected image during a second time period.

7. The system of Claim 1 wherein the excitation light and the reference light are emitted simultaneously such that respective images are detected by a color-sensitive image sensor, a blue channel detecting the fluorescence image and a red channel detecting the reference image.
- 5 8. The system of Claim 1 wherein the excitation light is in the range of 300 to 420 nm.
9. The system of Claim 1 wherein the light source further comprises a reference light source having a wavelength in a red or infrared range.
- 10 10. The system of Claim 1 wherein the optical guide comprises an optical fiber with a distally mounted lens.
11. The system of Claim 1 wherein the excitation light has an angular orientation that is the same as an angular orientation as the reference light.
12. The system of Claim 1 further comprising a reference illumination of a selected wavelength to provide a second reference image.
- 15 13. The system of Claim 12 further comprising a light source and a first filter wheel that filters light from the light source to provide a first reference illumination wavelength and a second illumination wavelength.
14. The system of Claim 1 further comprising a filter wheel having a filter that filters light from the light source to provide fluorescence illumination.

15. A method for imaging tissue fluorescence comprising:
detecting a fluorescence image of tissue and a reference image of the
tissue; and
processing said fluorescence image together with said reference image to
5 produce an output image of the tissue.
16. The method of Claim 15 further comprising providing an arc lamp light source.
17. The method of Claim 16 further comprising pulsing the arc lamp current source.
18. The method of Claim 15 further comprising sequentially the images with a
monochromatic image sensor.
- 10 19. The method of Claim 15 further comprising detecting images with a color-
sensitive image sensor, the sensor having a blue channel detecting an
autofluorescence image and a red channel detecting the reference image.
20. The method of Claim 15 further comprising coupling excitation light and a
reference light to an optical fiber such that a variation in a normalized intensity
15 of the reference light and a normalized intensity of the excitation light is less
than 20% at any point in a wavefront along the optical path between a combiner
that combines the excitation light and the reference light.
21. A method for imaging tissue fluorescence comprising:
providing excitation light having a wavelength in a range of 300 nm to
20 420 nm;
providing a reference light;

combining said excitation light and said reference light onto a common optical path such that an intensity of the excitation light varies less than 20% relative to a normalized intensity of the reference light at any point along the optical path;

5 detecting a fluorescence image of the tissue due to the said excitation light and a reference image of the tissue due to reflected reference light with an imaging sensor at a distal end of an endoscopic probe; and

 processing said fluorescence image and said reference image to produce an output image of the tissue.

- 10 22. The method of Claim 21 further comprising determining a ratio of the fluorescence image and the reference image to provide a processed image.
23. The method of Claim 21 further comprising adjusting the relative intensity or angular distribution of the reference light relative to the excitation light.
24. The method of Claim 21 further comprising detecting first and second reference
15 images having different wavelengths.
25. The method of Claim 24 further comprising detecting blue and red reference images and forming a third reference image from a pixel-by-pixel ratio of the blue and red images.

1/13




Figure 1




Figure 2a




Figure 2b

3/13





Figure 3



5/13




Figure 5

6/13




Figure 6a




Figure 6b




Figure 6c

7/13




Figure 7

8/13




Figure 9a




Figure 9b





Figure 9c



10/13










Figure 10b

11/13



12/13



13/13




Figure 13a




Figure 13b

INTERNATIONAL SEARCH REPORT

Intern. Appl. No.

PCT/US 00/01673

A. CLASSIFICATION OF SUBJECT MATTER
IPC 7 A61B5/00

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

IPC 7 A61B G01N

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X A	US 5 187 572 A (NAKAMURA KAZUNARI ET AL) 16 February 1993 (1993-02-16) column 15, line 37 -column 16, line 3 column 19, line 12 - line 31 column 3, line 65 -column 4, line 39 column 9, line 22 - line 28	15-18, 20,21 1-6, 8-11,14, 23
X A	US 5 769 792 A (HUNG JACLYN Y-C ET AL) 23 June 1998 (1998-06-23) column 4, line 24 - line 35 column 5, line 12 - line 63 column 7, line 1 - line 15 column 7, line 31 -column 8, line 14 -/-	15,18, 20,21,24 1,4,6-8, 11-14, 19,22, 24,25

☒ Further documents are listed in the continuation of box C.☒ Patent family members are listed in annex.

* Special categories of cited documents :

"A" document defining the general state of the art which is not considered to be of particular relevance

"E" earlier document but published on or after the international filing date

"L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)

"O" document referring to an oral disclosure, use, exhibition or other means

"P" document published prior to the international filing date but later than the priority date claimed

"T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention

"X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone

"Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.

"&" document member of the same patent family

Date of the actual completion of the international search

8 June 2000

Date of mailing of the international search report

16/06/2000

Name and mailing address of the ISA

European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040, Tx. 31 651 epo rd,
Fax: (+31-70) 340-3016

Authorized officer

Knüpling, M

INTERNATIONAL SEARCH REPORT

International Application No

PCT/US 00/01673

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT		
Category *	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 5 413 108 A (ALFANO ROBERT R) 9 May 1995 (1995-05-09)	15
Y	column 4, line 15 -column 5, line 4	1,4,6,8, 11,14
A		2,16,18
Y	DE 196 27 369 A (HEINRICH HERRMANN DR; SPECK UWE DIPL ING (DE)) 9 October 1997 (1997-10-09) column 2, line 30 -column 3, line 61	1,4,6,8, 11,14
A	US 4 449 535 A (RENAULT GUY) 22 May 1984 (1984-05-22) column 2, line 49 - line 63 column 4, line 66 -column 5, line 23	1,4, 7-11,15, 21,22

INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No

PCT/US 00/01673

Patent document cited in search report	Publication date	Patent family member(s)	Publication date
US 5187572 A	16-02-1993	JP 5084218 A DE 4136034 A	06-04-1993 07-05-1992
US 5769792 A	23-06-1998	US 5507287 A AT 140600 T CA 2042075 A DE 69212382 D EP 0512965 A JP 6054792 A	16-04-1996 15-08-1996 09-11-1992 29-08-1996 11-11-1992 01-03-1994
US 5413108 A	09-05-1995	NONE	
DE 19627369 A	09-10-1997	NONE	
US 4449535 A	22-05-1984	FR 2502783 A FR 2521727 A DE 3210593 A JP 1588724 C JP 2007653 B JP 57188242 A	01-10-1982 19-08-1983 28-10-1982 19-11-1990 20-02-1990 19-11-1982